

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO051525

Review By	yogesh	Review On	5/14/2025 3:39:03 PM
Supervise By	mohammad	Supervise On	5/16/2025 1:55:32 AM
SubDirectory	PO051525	HP Acquire Method	HP Processing Method PO051525

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO111056.D	14 May 2025 16:43		YP/AJ	Ok
2	I.BLK	I.BLK	PO111057.D	14 May 2025 17:01		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO111058.D	14 May 2025 17:19		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO111059.D	14 May 2025 17:38		YP/AJ	Ok,M
5	AR1660ICC500	AR1660ICC500	PO111060.D	14 May 2025 17:56		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO111061.D	14 May 2025 18:15		YP/AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PO111062.D	14 May 2025 18:33		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO111063.D	14 May 2025 18:51		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO111064.D	14 May 2025 19:10		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO111065.D	14 May 2025 19:28		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO111066.D	14 May 2025 19:46		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO111067.D	14 May 2025 20:05		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO111068.D	14 May 2025 20:23		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO111069.D	14 May 2025 20:41		YP/AJ	Ok
15	AR1248ICC1000	AR1248ICC1000	PO111070.D	14 May 2025 21:00		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO111071.D	14 May 2025 21:18		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO111072.D	14 May 2025 21:37		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO111073.D	14 May 2025 21:54		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO051525

Review By	yogesh	Review On	5/14/2025 3:39:03 PM
Supervise By	mohammad	Supervise On	5/16/2025 1:55:32 AM
SubDirectory	PO051525	HP Acquire Method	HP Processing Method PO051525
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

19	AR1248ICC050	AR1248ICC050	PO111074.D	14 May 2025 22:12		YP/AJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PO111075.D	14 May 2025 22:31		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO111076.D	14 May 2025 22:49		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO111077.D	14 May 2025 23:06		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO111078.D	14 May 2025 23:25		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO111079.D	14 May 2025 23:43		YP/AJ	Ok
25	AR1262ICC500	AR1262ICC500	PO111080.D	15 May 2025 00:01		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO111081.D	15 May 2025 00:19		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO111082.D	15 May 2025 00:37		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO111083.D	15 May 2025 00:56		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO111084.D	15 May 2025 01:14		YP/AJ	Ok
30	AR1268ICC050	AR1268ICC050	PO111085.D	15 May 2025 01:31		YP/AJ	Ok
31	PO051525ICV500	ICVPO051525	PO111086.D	15 May 2025 01:50		YP/AJ	Ok
32	AR1242ICV500	ICVPO051525AR1242	PO111087.D	15 May 2025 02:26		YP/AJ	Ok
33	AR1248ICV500	ICVPO051525AR1248	PO111088.D	15 May 2025 03:03		YP/AJ	Ok
34	AR1254ICV500	ICVPO051525AR1254	PO111089.D	15 May 2025 03:40		YP/AJ	Ok
35	AR1268ICV500	ICVPO051525AR1268	PO111090.D	15 May 2025 04:17		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO060425

Review By	yogesh	Review On	6/4/2025 10:15:40 AM
Supervise By	mohammad	Supervise On	6/6/2025 2:01:29 AM
SubDirectory	PO060425	HP Acquire Method	HP Processing Method PO051525

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO111452.D	04 Jun 2025 08:39		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO111453.D	04 Jun 2025 08:57		YP/AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PO111454.D	04 Jun 2025 09:18		YP/AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PO111455.D	04 Jun 2025 09:36		YP/AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PO111456.D	04 Jun 2025 09:54		YP/AJ	Ok,M
6	I.BLK	I.BLK	PO111457.D	04 Jun 2025 10:13		YP/AJ	Ok,M
7	Q2175-01	32525	PO111458.D	04 Jun 2025 10:31	AR1254 hits	YP/AJ	Ok,M
8	Q2176-01	TP-46	PO111459.D	04 Jun 2025 10:49		YP/AJ	Ok,M
9	Q2176-02	TP-56	PO111460.D	04 Jun 2025 11:08		YP/AJ	Ok,M
10	Q2176-03	TP-25	PO111461.D	04 Jun 2025 11:26	AR1260 Hit	YP/AJ	Ok,M
11	Q2175-06RE	EGR-LIQUIDRE	PO111462.D	04 Jun 2025 11:50	All surrogate low	YP/AJ	Confirms
12	Q2171-02	5-28-25-B	PO111463.D	04 Jun 2025 12:57	AR1242 Hit	YP/AJ	Ok,M
13	Q2171-03	51925	PO111464.D	04 Jun 2025 13:15	AR1242 Hit	YP/AJ	Ok,M
14	Q2193-01	VNJ259-1	PO111465.D	04 Jun 2025 13:33	AR1242 Hit	YP/AJ	Ok,M
15	Q2193-02	VNJ259-2	PO111466.D	04 Jun 2025 13:51	AR1242 Hit	YP/AJ	Ok,M
16	Q2193-03	VNJ259-3	PO111467.D	04 Jun 2025 14:09	AR1248 Hit	YP/AJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PO111468.D	04 Jun 2025 15:23		YP/AJ	Ok,M
18	AR1242CCC500	AR1242CCC500	PO111469.D	04 Jun 2025 15:41		YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO060425

Review By	yogesh	Review On	6/4/2025 10:15:40 AM
Supervise By	mohammad	Supervise On	6/6/2025 2:01:29 AM
SubDirectory	PO060425	HP Acquire Method	HP Processing Method PO051525
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PO111470.D	04 Jun 2025 15:59		YP/AJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PO111471.D	04 Jun 2025 16:18		YP/AJ	Ok,M
21	I.BLK	I.BLK	PO111472.D	04 Jun 2025 16:35		YP/AJ	Ok,M
22	Q2198-03	B-207-SB02	PO111473.D	04 Jun 2025 16:52		YP/AJ	Ok,M
23	Q2199-01	ETGI-343	PO111474.D	04 Jun 2025 17:10		YP/AJ	Ok,M
24	Q2199-03	VNJ-231	PO111475.D	04 Jun 2025 17:28		YP/AJ	Ok,M
25	Q2199-05	72-11978	PO111476.D	04 Jun 2025 17:45		YP/AJ	Ok,M
26	AR1660CCC500	AR1660CCC500	PO111477.D	04 Jun 2025 18:57		YP/AJ	Ok,M
27	AR1242CCC500	AR1242CCC500	PO111478.D	04 Jun 2025 19:33		YP/AJ	Ok,M
28	AR1248CCC500	AR1248CCC500	PO111479.D	04 Jun 2025 19:50		YP/AJ	Ok,M
29	AR1254CCC500	AR1254CCC500	PO111480.D	04 Jun 2025 20:08		YP/AJ	Ok,M
30	I.BLK	I.BLK	PO111481.D	04 Jun 2025 20:27		YP/AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP072162.D	19 May 2025 09:09		YP\AJ	Ok
2	I.BLK	I.BLK	PP072163.D	19 May 2025 09:26		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP072164.D	19 May 2025 09:42		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP072165.D	19 May 2025 09:59		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP072166.D	19 May 2025 10:15		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP072167.D	19 May 2025 10:32		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP072168.D	19 May 2025 10:48		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP072169.D	19 May 2025 11:05		YP\AJ	Ok
9	AR1232ICC1000	AR1232ICC1000	PP072170.D	19 May 2025 11:21		YP\AJ	Ok
10	AR1232ICC750	AR1232ICC750	PP072171.D	19 May 2025 11:38		YP\AJ	Ok
11	AR1232ICC500	AR1232ICC500	PP072172.D	19 May 2025 11:54		YP\AJ	Ok
12	AR1232ICC250	AR1232ICC250	PP072173.D	19 May 2025 12:10		YP\AJ	Ok
13	AR1232ICC050	AR1232ICC050	PP072174.D	19 May 2025 12:27		YP\AJ	Ok,M
14	AR1242ICC1000	AR1242ICC1000	PP072175.D	19 May 2025 12:43		YP\AJ	Ok
15	AR1242ICC750	AR1242ICC750	PP072176.D	19 May 2025 12:59		YP\AJ	Ok
16	AR1242ICC500	AR1242ICC500	PP072177.D	19 May 2025 13:15		YP\AJ	Ok
17	AR1242ICC250	AR1242ICC250	PP072178.D	19 May 2025 13:32		YP\AJ	Ok
18	AR1242ICC050	AR1242ICC050	PP072179.D	19 May 2025 13:48		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

19	AR1248ICC1000	AR1248ICC1000	PP072180.D	19 May 2025 14:04		YPIAJ	Ok
20	AR1248ICC750	AR1248ICC750	PP072181.D	19 May 2025 14:20		YPIAJ	Ok
21	AR1248ICC500	AR1248ICC500	PP072182.D	19 May 2025 14:37		YPIAJ	Ok
22	AR1248ICC250	AR1248ICC250	PP072183.D	19 May 2025 14:53		YPIAJ	Ok
23	AR1248ICC050	AR1248ICC050	PP072184.D	19 May 2025 15:25		YPIAJ	Ok,M
24	AR1254ICC1000	AR1254ICC1000	PP072185.D	19 May 2025 15:42		YPIAJ	Ok
25	AR1254ICC750	AR1254ICC750	PP072186.D	19 May 2025 15:58		YPIAJ	Ok
26	AR1254ICC500	AR1254ICC500	PP072187.D	19 May 2025 16:14		YPIAJ	Ok
27	AR1254ICC250	AR1254ICC250	PP072188.D	19 May 2025 16:31		YPIAJ	Ok
28	AR1254ICC050	AR1254ICC050	PP072189.D	19 May 2025 16:47		YPIAJ	Ok,M
29	AR1262ICC500	AR1262ICC500	PP072190.D	19 May 2025 17:03		YPIAJ	Ok
30	AR1268ICC1000	AR1268ICC1000	PP072191.D	19 May 2025 17:20		YPIAJ	Ok
31	AR1268ICC750	AR1268ICC750	PP072192.D	19 May 2025 17:36		YPIAJ	Ok
32	AR1268ICC500	AR1268ICC500	PP072193.D	19 May 2025 17:53		YPIAJ	Ok
33	AR1268ICC250	AR1268ICC250	PP072194.D	19 May 2025 18:09		YPIAJ	Ok
34	AR1268ICC050	AR1268ICC050	PP072195.D	19 May 2025 18:25		YPIAJ	Ok,M
35	PP051925ICV500	ICVPP051925	PP072196.D	19 May 2025 18:42		YPIAJ	Ok
36	AR1232ICV500	ICVPP051925AR1232	PP072197.D	19 May 2025 19:14		YPIAJ	Ok
37	AR1242ICV500	ICVPP051925AR1242	PP072198.D	19 May 2025 19:47		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

38	AR1248ICV500	ICVPP051925AR1248	PP072199.D	19 May 2025 20:19		YPIAJ	Ok
39	AR1254ICV500	ICVPP051925AR1254	PP072200.D	19 May 2025 20:52		YPIAJ	Ok
40	AR1268ICV500	ICVPP051925AR1268	PP072201.D	19 May 2025 21:25		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP060325

Review By	yogesh	Review On	6/3/2025 11:21:30 AM
Supervise By	mohammad	Supervise On	6/5/2025 1:35:08 AM
SubDirectory	PP060325	HP Acquire Method	HP Processing Method PP051925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP072552.D	03 Jun 2025 10:06		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP072553.D	03 Jun 2025 10:22		YP\AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP072554.D	03 Jun 2025 10:39		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP072555.D	03 Jun 2025 10:55		YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP072556.D	03 Jun 2025 11:11		YP\AJ	Ok,M
6	I.BLK	I.BLK	PP072557.D	03 Jun 2025 11:27		YP\AJ	Ok,M
7	PB168252BL	PB168252BL	PP072558.D	03 Jun 2025 13:28		YP\AJ	Ok,M
8	PB168252BS	PB168252BS	PP072559.D	03 Jun 2025 13:44	Recovery fail	YP\AJ	Not Ok
9	Q2168-01	SAN-A1-A3	PP072560.D	03 Jun 2025 14:00	AR1242 Hit,Need Dilution 50X	YP\AJ	Dilution
10	Q2168-05	SAN-B1-B3	PP072561.D	03 Jun 2025 14:16	AR1242 Hit,Need Dilution 50X	YP\AJ	Dilution
11	Q2168-09	SAN-C1-C2	PP072562.D	03 Jun 2025 14:33	AR1242 Hit,Need Dilution 100X	YP\AJ	Dilution
12	Q2175-01	32525	PP072563.D	03 Jun 2025 14:49	AR1242 Hit,carry over	YP\AJ	Not Ok
13	Q2176-01	TP-46	PP072564.D	03 Jun 2025 15:05	AR1242 Hit,carry over	YP\AJ	Not Ok
14	Q2176-02	TP-56	PP072565.D	03 Jun 2025 15:22	AR1242 Hit,carry over	YP\AJ	Not Ok
15	Q2176-03	TP-25	PP072566.D	03 Jun 2025 15:38	AR1260 Hit; possible carryover	YP\AJ	Not Ok
16	Q2176-04	TP-26	PP072567.D	03 Jun 2025 15:54		YP\AJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PP072568.D	03 Jun 2025 16:59		YP\AJ	Ok,M
18	AR1242CCC500	AR1242CCC500	PP072569.D	03 Jun 2025 17:32		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP060325

Review By	yogesh	Review On	6/3/2025 11:21:30 AM
Supervise By	mohammad	Supervise On	6/5/2025 1:35:08 AM
SubDirectory	PP060325	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP072570.D	03 Jun 2025 17:48		YPIAJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP072571.D	03 Jun 2025 18:05		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP072572.D	03 Jun 2025 18:21		YPIAJ	Ok,M
22	Q2176-05	TP-28	PP072573.D	03 Jun 2025 18:37		YPIAJ	Ok
23	Q2176-05MS	TP-28MS	PP072574.D	03 Jun 2025 18:53		YPIAJ	Ok,M
24	Q2176-05MSD	TP-28MSD	PP072575.D	03 Jun 2025 19:10		YPIAJ	Ok
25	Q2176-06	TP-27	PP072576.D	03 Jun 2025 19:26		YPIAJ	Ok,M
26	Q2176-07	TP-31	PP072577.D	03 Jun 2025 19:42		YPIAJ	Ok
27	Q2176-08	TP-65	PP072578.D	03 Jun 2025 19:59		YPIAJ	Ok
28	Q2177-02	B-187-SB01	PP072579.D	03 Jun 2025 20:15		YPIAJ	Ok,M
29	Q2177-04	B-187-SB02	PP072580.D	03 Jun 2025 20:31		YPIAJ	Ok,M
30	Q2177-06	B-202-SB01	PP072581.D	03 Jun 2025 20:48		YPIAJ	Ok,M
31	Q2178-01	RT2929	PP072582.D	03 Jun 2025 21:04		YPIAJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PP072583.D	03 Jun 2025 22:25		YPIAJ	Ok,M
33	AR1242CCC500	AR1242CCC500	PP072584.D	03 Jun 2025 22:58		YPIAJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PP072585.D	03 Jun 2025 23:14		YPIAJ	Ok,M
35	AR1254CCC500	AR1254CCC500	PP072586.D	03 Jun 2025 23:30		YPIAJ	Ok,M
36	I.BLK	I.BLK	PP072587.D	03 Jun 2025 23:47		YPIAJ	Ok,M
37	Q2185-01	TP02-MHB-WC	PP072588.D	04 Jun 2025 00:03		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP060325

Review By	yogesh	Review On	6/3/2025 11:21:30 AM
Supervise By	mohammad	Supervise On	6/5/2025 1:35:08 AM
SubDirectory	PP060325	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q2185-05	TP01-MHA-WC	PP072589.D	04 Jun 2025 00:19		YPIAJ	Ok,M
39	PB168257BL	PB168257BL	PP072590.D	04 Jun 2025 00:35		YPIAJ	Ok,M
40	PB168257BS	PB168257BS	PP072591.D	04 Jun 2025 00:52	AR1016 recovery fail	YPIAJ	Not Ok
41	PB168257BSD	PB168257BSD	PP072592.D	04 Jun 2025 01:08	AR1016 recovery fail	YPIAJ	Not Ok
42	Q2165-03	TRE-25-63-64	PP072593.D	04 Jun 2025 01:24	AR1254 Hit	YPIAJ	Ok,M
43	Q2165-06	TRE-25-65-66	PP072594.D	04 Jun 2025 01:40		YPIAJ	Ok,M
44	Q2169-01	303-PPR-1	PP072595.D	04 Jun 2025 01:57	AR1254 Hit	YPIAJ	Ok,M
45	Q2169-03	303-PPR-2	PP072596.D	04 Jun 2025 02:13	AR1254 Hit	YPIAJ	Ok,M
46	Q2175-04	52525-A-B-COMP	PP072597.D	04 Jun 2025 02:29		YPIAJ	Ok,M
47	AR1660CCC500	AR1660CCC500	PP072598.D	04 Jun 2025 03:51		YPIAJ	Ok,M
48	AR1242CCC500	AR1242CCC500	PP072599.D	04 Jun 2025 04:23		YPIAJ	Ok,M
49	AR1248CCC500	AR1248CCC500	PP072600.D	04 Jun 2025 04:40		YPIAJ	Ok,M
50	AR1254CCC500	AR1254CCC500	PP072601.D	04 Jun 2025 04:56		YPIAJ	Ok,M
51	I.BLK	I.BLK	PP072602.D	04 Jun 2025 05:12		YPIAJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP060425

Review By	yogesh	Review On	6/4/2025 10:14:35 AM
Supervise By	mohammad	Supervise On	6/6/2025 9:37:42 AM
SubDirectory	PP060425	HP Acquire Method	HP Processing Method PP051925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP072603.D	04 Jun 2025 08:34		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP072604.D	04 Jun 2025 08:50		YP\AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP072605.D	04 Jun 2025 09:11		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP072606.D	04 Jun 2025 09:27		YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP072607.D	04 Jun 2025 09:43		YP\AJ	Ok,M
6	I.BLK	I.BLK	PP072608.D	04 Jun 2025 10:00		YP\AJ	Ok,M
7	PB168257BS	PB168257BS	PP072609.D	04 Jun 2025 10:16		YP\AJ	Ok,M
8	PB168257BSD	PB168257BSD	PP072610.D	04 Jun 2025 10:32		YP\AJ	Ok,M
9	Q2168-01DL	SAN-A1-A3DL	PP072611.D	04 Jun 2025 10:48	AR1242 Hit	YP\AJ	Ok
10	Q2168-05DL	SAN-B1-B3DL	PP072612.D	04 Jun 2025 11:04	AR1242 Hit	YP\AJ	Ok
11	Q2168-09DL	SAN-C1-C2DL	PP072613.D	04 Jun 2025 11:21	AR1242 Hit	YP\AJ	Ok
12	PB168252BS	PB168252BS	PP072614.D	04 Jun 2025 11:37		YP\AJ	Ok,M
13	PB168275BL	PB168275BL	PP072615.D	04 Jun 2025 12:55		YP\AJ	Ok
14	PB168275BS	PB168275BS	PP072616.D	04 Jun 2025 13:12	AR1016 fail for recovery	YP\AJ	Not Ok
15	Q2167-01	52225	PP072617.D	04 Jun 2025 13:28	Surrogate fail, AR1242 Hit	YP\AJ	ReRun
16	Q2171-01	5-28-25-A	PP072618.D	04 Jun 2025 13:45	AR1242 Hit,Need Dilution, DCB high 2nd column	YP\AJ	Dilution
17	AR1660CCC500	AR1660CCC500	PP072619.D	04 Jun 2025 14:51		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP060425

Review By	yogesh	Review On	6/4/2025 10:14:35 AM
Supervise By	mohammad	Supervise On	6/6/2025 9:37:42 AM
SubDirectory	PP060425	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PP072620.D	04 Jun 2025 15:07		YPIAJ	Ok,M
19	AR1248CCC500	AR1248CCC500	PP072621.D	04 Jun 2025 15:23		YPIAJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP072622.D	04 Jun 2025 15:40		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP072623.D	04 Jun 2025 15:56		YPIAJ	Ok
22	PB168288BL	PB168288BL	PP072624.D	04 Jun 2025 16:13		YPIAJ	Ok,M
23	PB168288BS	PB168288BS	PP072625.D	04 Jun 2025 16:29	AR1016 fail for recovery	YPIAJ	Not Ok
24	Q2182-01	OR-03-06022025	PP072626.D	04 Jun 2025 16:45		YPIAJ	Ok,M
25	Q2192-01	SB-1	PP072627.D	04 Jun 2025 17:02	AR1260 + 1242 Hit , need 25X dilution	YPIAJ	Dilution
26	Q2194-01	COMP-12	PP072628.D	04 Jun 2025 17:18		YPIAJ	Ok
27	Q2194-03	COMP-13	PP072629.D	04 Jun 2025 17:34		YPIAJ	Ok
28	Q2195-01	OK-01-060325	PP072630.D	04 Jun 2025 17:51		YPIAJ	Ok,M
29	Q2195-01MS	OK-01-060325MS	PP072631.D	04 Jun 2025 18:07		YPIAJ	Ok,M
30	Q2195-01MSD	OK-01-060325MSD	PP072632.D	04 Jun 2025 18:23		YPIAJ	Ok,M
31	Q2198-01	B-202-SB02	PP072633.D	04 Jun 2025 18:40		YPIAJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PP072634.D	04 Jun 2025 20:01	AR1016 - 1 low in 1st column , AR1260 - 2,4 low in 1st column	YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PP072635.D	04 Jun 2025 20:18	AR1242 - 5 Low in 1st column	YPIAJ	Ok
34	AR1248CCC500	AR1248CCC500	PP072636.D	04 Jun 2025 20:34		YPIAJ	Ok,M
35	AR1254CCC500	AR1254CCC500	PP072637.D	04 Jun 2025 20:50	AR1254 - 2,4 low in 1st column	YPIAJ	Ok,M
36	I.BLK	I.BLK	PP072638.D	04 Jun 2025 21:07		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP060425

Review By	yogesh	Review On	6/4/2025 10:14:35 AM
Supervise By	mohammad	Supervise On	6/6/2025 9:37:42 AM
SubDirectory	PP060425	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	Q2175-09	52725	PP072639.D	04 Jun 2025 21:23	AR1242 hit, DCB low in both column & TCMX low in 1st column	YPIAJ	ReRun
38	Q2171-01DL	5-28-25-ADL	PP072640.D	04 Jun 2025 21:39	AR1242 Hit	YPIAJ	Ok,M
39	AR1660CCC500	AR1660CCC500	PP072641.D	04 Jun 2025 22:45		YPIAJ	Ok,M
40	AR1242CCC500	AR1242CCC500	PP072642.D	04 Jun 2025 23:17		YPIAJ	Ok,M
41	AR1248CCC500	AR1248CCC500	PP072643.D	04 Jun 2025 23:34		YPIAJ	Ok,M
42	AR1254CCC500	AR1254CCC500	PP072644.D	04 Jun 2025 23:50		YPIAJ	Ok,M
43	I.BLK	I.BLK	PP072645.D	05 Jun 2025 00:06		YPIAJ	Ok,M

M : Manual Integration