

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL01 Date Analyzed: 06/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL095880.D Time Analyzed: 09:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.732	6.634	6.834	54.330	50.000	8.7
4,4'-DDE	6.223	6.124	6.324	52.600	50.000	5.2
4,4'-DDT	7.048	6.950	7.150	50.810	50.000	1.6
Aldrin	5.296	5.197	5.397	55.700	50.000	11.4
alpha-BHC	4.024	3.925	4.125	55.670	50.000	11.3
alpha-Chlordane	6.053	5.954	6.154	54.590	50.000	9.2
beta-BHC	4.543	4.444	4.644	54.960	50.000	9.9
Decachlorobiphenyl	9.099	9.003	9.203	50.410	50.000	0.8
delta-BHC	4.792	4.692	4.892	54.900	50.000	9.8
Dieldrin	6.374	6.275	6.475	54.900	50.000	9.8
Endosulfan I	6.100	6.002	6.202	54.550	50.000	9.1
Endosulfan II	6.814	6.715	6.915	49.230	50.000	-1.5
Endosulfan sulfate	7.176	7.078	7.278	52.640	50.000	5.3
Endrin	6.601	6.502	6.702	51.610	50.000	3.2
Endrin aldehyde	6.942	6.844	7.044	53.510	50.000	7.0
Endrin ketone	7.658	7.560	7.760	53.170	50.000	6.3
gamma-BHC (Lindane)	4.356	4.257	4.457	55.600	50.000	11.2
gamma-Chlordane	5.972	5.873	6.073	55.720	50.000	11.4
Heptachlor	4.953	4.855	5.055	54.350	50.000	8.7
Heptachlor epoxide	5.717	5.618	5.818	55.510	50.000	11.0
Methoxychlor	7.521	7.423	7.623	51.720	50.000	3.4
Tetrachloro-m-xylene	3.573	3.473	3.673	55.690	50.000	11.4

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL01 Date Analyzed: 06/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL095880.D Time Analyzed: 09:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.923	5.826	6.026	54.870	50.000	9.7
4,4'-DDE	5.372	5.274	5.474	53.030	50.000	6.1
4,4'-DDT	6.176	6.080	6.280	50.640	50.000	1.3
Aldrin	4.366	4.268	4.468	57.020	50.000	14.0
alpha-BHC	3.395	3.296	3.496	56.700	50.000	13.4
alpha-Chlordane	5.185	5.087	5.287	55.530	50.000	11.1
beta-BHC	4.024	3.926	4.126	55.470	50.000	10.9
Decachlorobiphenyl	8.059	7.964	8.164	51.100	50.000	2.2
delta-BHC	4.260	4.161	4.361	55.570	50.000	11.1
Dieldrin	5.505	5.408	5.608	55.500	50.000	11.0
Endosulfan I	5.240	5.143	5.343	50.760	50.000	1.5
Endosulfan II	6.069	5.973	6.173	54.320	50.000	8.6
Endosulfan sulfate	6.471	6.375	6.575	53.440	50.000	6.9
Endrin	5.778	5.683	5.883	51.800	50.000	3.6
Endrin aldehyde	6.248	6.151	6.351	54.090	50.000	8.2
Endrin ketone	6.977	6.880	7.080	53.690	50.000	7.4
gamma-BHC (Lindane)	3.729	3.630	3.830	56.170	50.000	12.3
gamma-Chlordane	5.120	5.023	5.223	55.590	50.000	11.2
Heptachlor	4.082	3.983	4.183	55.200	50.000	10.4
Heptachlor epoxide	4.868	4.771	4.971	55.170	50.000	10.3
Methoxychlor	6.747	6.651	6.851	49.000	50.000	-2.0
Tetrachloro-m-xylene	2.886	2.786	2.986	54.700	50.000	9.4