

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL02 Date Analyzed: 06/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL095897.D Time Analyzed: 17:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.733	6.634	6.834	46.270	50.000	-7.5
4,4'-DDE	6.222	6.124	6.324	46.320	50.000	-7.4
4,4'-DDT	7.047	6.950	7.150	44.000	50.000	-12.0
Aldrin	5.296	5.197	5.397	48.690	50.000	-2.6
alpha-BHC	4.024	3.925	4.125	50.060	50.000	0.1
alpha-Chlordane	6.052	5.954	6.154	47.030	50.000	-5.9
beta-BHC	4.543	4.444	4.644	49.130	50.000	-1.7
Decachlorobiphenyl	9.098	9.003	9.203	43.980	50.000	-12.0
delta-BHC	4.791	4.692	4.892	49.560	50.000	-0.9
Dieldrin	6.373	6.275	6.475	47.010	50.000	-6.0
Endosulfan I	6.100	6.002	6.202	47.200	50.000	-5.6
Endosulfan II	6.814	6.715	6.915	42.140	50.000	-15.7
Endosulfan sulfate	7.176	7.078	7.278	45.150	50.000	-9.7
Endrin	6.601	6.502	6.702	44.600	50.000	-10.8
Endrin aldehyde	6.942	6.844	7.044	46.040	50.000	-7.9
Endrin ketone	7.657	7.560	7.760	46.370	50.000	-7.3
gamma-BHC (Lindane)	4.356	4.257	4.457	49.830	50.000	-0.3
gamma-Chlordane	5.971	5.873	6.073	47.550	50.000	-4.9
Heptachlor	4.953	4.855	5.055	48.950	50.000	-2.1
Heptachlor epoxide	5.717	5.618	5.818	48.650	50.000	-2.7
Methoxychlor	7.520	7.423	7.623	44.540	50.000	-10.9
Tetrachloro-m-xylene	3.572	3.473	3.673	50.900	50.000	1.8

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL02 Date Analyzed: 06/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL095897.D Time Analyzed: 17:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.922	5.826	6.026	47.940	50.000	-4.1
4,4'-DDE	5.371	5.274	5.474	45.770	50.000	-8.5
4,4'-DDT	6.177	6.080	6.280	41.460	50.000	-17.1
Aldrin	4.364	4.268	4.468	51.850	50.000	3.7
alpha-BHC	3.395	3.296	3.496	53.460	50.000	6.9
alpha-Chlordane	5.184	5.087	5.287	49.960	50.000	-0.1
beta-BHC	4.024	3.926	4.126	51.510	50.000	3.0
Decachlorobiphenyl	8.059	7.964	8.164	42.370	50.000	-15.3
delta-BHC	4.259	4.161	4.361	52.410	50.000	4.8
Dieldrin	5.505	5.408	5.608	48.900	50.000	-2.2
Endosulfan I	5.240	5.143	5.343	45.990	50.000	-8.0
Endosulfan II	6.070	5.973	6.173	46.740	50.000	-6.5
Endosulfan sulfate	6.471	6.375	6.575	44.900	50.000	-10.2
Endrin	5.779	5.683	5.883	45.120	50.000	-9.8
Endrin aldehyde	6.248	6.151	6.351	46.350	50.000	-7.3
Endrin ketone	6.977	6.880	7.080	43.910	50.000	-12.2
gamma-BHC (Lindane)	3.729	3.630	3.830	52.270	50.000	4.5
gamma-Chlordane	5.120	5.023	5.223	49.880	50.000	-0.2
Heptachlor	4.081	3.983	4.183	51.390	50.000	2.8
Heptachlor epoxide	4.868	4.771	4.971	49.980	50.000	0.0
Methoxychlor	6.747	6.651	6.851	40.990	50.000	-18.0
Tetrachloro-m-xylene	2.886	2.786	2.986	51.450	50.000	2.9