

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF052025

Review By	Rahul	Review On	5/21/2025 2:52:20 PM
Supervise By	Jagrut	Supervise On	5/21/2025 4:44:06 PM
SubDirectory	BF052025	HP Acquire Method	BNA_F
		HP Processing Method	bf052025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12665,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142465.D	20 May 2025 11:13		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142466.D	20 May 2025 11:41	Fresh Calibration Required	RC/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BF142467.D	20 May 2025 12:10		RC/JU	Ok
4	SSTDICC005	SSTDICC005	BF142468.D	20 May 2025 12:38	Compound #32,54,85 removed from 5ppm	RC/JU	Ok
5	SSTDICC010	SSTDICC010	BF142469.D	20 May 2025 13:07		RC/JU	Ok,M
6	SSTDICC020	SSTDICC020	BF142470.D	20 May 2025 13:36		RC/JU	Ok
7	SSTDICCC040	SSTDICCC040	BF142471.D	20 May 2025 14:05	This calibration is good for both the methods, 8270E DOD and 625.1.	RC/JU	Ok
8	SSTDICC050	SSTDICC050	BF142472.D	20 May 2025 14:34		RC/JU	Ok
9	SSTDICC060	SSTDICC060	BF142473.D	20 May 2025 15:03		RC/JU	Ok
10	SSTDICC080	SSTDICC080	BF142474.D	20 May 2025 15:31		RC/JU	Ok
11	SSTDICV040	ICVBF052025	BF142475.D	20 May 2025 16:31		RC/JU	Ok
12	PB168067TB	PB168067TB	BF142476.D	20 May 2025 17:29		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF060425

Review By	Rahul	Review On	6/5/2025 11:49:57 AM
Supervise By	Jagrut	Supervise On	6/5/2025 1:15:58 PM
SubDirectory	BF060425	HP Acquire Method	BNA_F
		HP Processing Method	bf052025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12667,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142601.D	04 Jun 2025 11:59		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142602.D	04 Jun 2025 12:27		RC/JU	Ok,M
3	PB168234BL	PB168234BL	BF142603.D	04 Jun 2025 13:05		RC/JU	Ok
4	Q2159-01	TP05-MHO-WC	BF142604.D	04 Jun 2025 13:39		RC/JU	Ok
5	Q2159-01MS	TP05-MHO-WCMS	BF142605.D	04 Jun 2025 14:08		RC/JU	Ok,M
6	Q2159-01MSD	TP05-MHO-WCMSD	BF142606.D	04 Jun 2025 14:38		RC/JU	Ok
7	Q2160-05	TP05-MHH-WC	BF142607.D	04 Jun 2025 15:17		RC/JU	Ok
8	Q2172-01	TP06-MHQ	BF142608.D	04 Jun 2025 15:47		RC/JU	Ok
9	Q2160-01	TP04-MHG-WC	BF142609.D	04 Jun 2025 16:17		RC/JU	Ok
10	Q2159-04	TP05-MHO-WC	BF142610.D	04 Jun 2025 16:47		RC/JU	Ok
11	Q2160-04	TP04-MHG-WC	BF142611.D	04 Jun 2025 17:17		RC/JU	Ok
12	Q2160-08	TP05-MHH-WC	BF142612.D	04 Jun 2025 17:48		RC/JU	Ok
13	Q2173-06	OR-400-CF-402B-COM	BF142613.D	04 Jun 2025 18:18		RC/JU	Ok
14	Q2173-12	OR-400-CF-402B-COM	BF142614.D	04 Jun 2025 18:47		RC/JU	Ok
15	Q2173-18	OR-400-CF-402B-COM	BF142615.D	04 Jun 2025 19:17		RC/JU	Ok
16	Q2173-07	OR-400-CF-402B-COM	BF142616.D	04 Jun 2025 19:47		RC/JU	Ok,M
17	Q2173-13	OR-400-CF-402B-COM	BF142617.D	04 Jun 2025 20:17		RC/JU	Ok,M
18	Q2173-01	OR-400-CF-402B-COM	BF142618.D	04 Jun 2025 20:47		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF060425

Review By	Rahul	Review On	6/5/2025 11:49:57 AM		
Supervise By	Jagrut	Supervise On	6/5/2025 1:15:58 PM		
SubDirectory	BF060425	HP Acquire Method	BNA_F	HP Processing Method	bf052025
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791				
CCC	SP6787				
Internal Standard/PEM	S12667,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2172-04	TP06-MHQ	BF142619.D	04 Jun 2025 21:17		RC/JU	Ok
20	Q2185-04	TP02-MHB-WC	BF142620.D	04 Jun 2025 21:46		RC/JU	Ok
21	Q2185-08	TP01-MHA-WC	BF142621.D	04 Jun 2025 22:16		RC/JU	Ok
22	Q2182-01	OR-03-06022025	BF142622.D	04 Jun 2025 22:45		RC/JU	Ok,M
23	Q2178-01	RT2929	BF142623.D	04 Jun 2025 23:15	Analyze with further 10X Dilution first & then decide	RC/JU	Dilution

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF060625

Review By	Rahul	Review On	6/9/2025 11:32:18 AM		
Supervise By	Jagrut	Supervise On	6/9/2025 12:07:45 PM		
SubDirectory	BF060625	HP Acquire Method	BNA_F	HP Processing Method	bf052025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12667,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142639.D	06 Jun 2025 11:01		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142640.D	06 Jun 2025 11:36		RC/JU	Ok
3	PB168259BL	PB168259BL	BF142641.D	06 Jun 2025 12:06		RC/JU	Ok
4	Q2185-01	TP02-MHB-WC	BF142642.D	06 Jun 2025 12:54		RC/JU	Ok
5	Q2207-37	BU-703-COMP-05	BF142643.D	06 Jun 2025 13:24		RC/JU	Ok
6	Q2207-37MS	BU-703-COMP-05MS	BF142644.D	06 Jun 2025 13:54		RC/JU	Ok
7	Q2207-37MSD	BU-703-COMP-05MSD	BF142645.D	06 Jun 2025 14:25		RC/JU	Ok
8	Q2219-01	72-11995	BF142646.D	06 Jun 2025 14:58		RC/JU	Ok,M
9	Q2214-01	RBR251425	BF142647.D	06 Jun 2025 15:29		RC/JU	Ok
10	Q2208-19	BU-703-COMP-08	BF142648.D	06 Jun 2025 16:00		RC/JU	Ok
11	Q2207-28	BU-703-COMP-04	BF142649.D	06 Jun 2025 16:30		RC/JU	Ok
12	Q2176-07	TP-31	BF142650.D	06 Jun 2025 17:01		RC/JU	Ok
13	Q2198-01	B-202-SB02	BF142651.D	06 Jun 2025 17:31		RC/JU	Ok
14	Q2177-04	B-187-SB02	BF142652.D	06 Jun 2025 18:01		RC/JU	Ok
15	Q2198-05	B-202-GW01	BF142653.D	06 Jun 2025 18:31		RC/JU	Ok
16	Q2207-10	BU-703-COMP-02	BF142654.D	06 Jun 2025 19:02		RC/JU	Ok
17	Q2207-19	BU-703-COMP-03	BF142655.D	06 Jun 2025 19:32		RC/JU	Ok
18	Q2208-10	BU-703-COMP-07	BF142656.D	06 Jun 2025 20:02		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF060625

Review By	Rahul	Review On	6/9/2025 11:32:18 AM		
Supervise By	Jagrut	Supervise On	6/9/2025 12:07:45 PM		
SubDirectory	BF060625	HP Acquire Method	BNA_F	HP Processing Method	bf052025
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791				
CCC	SP6787				
Internal Standard/PEM	S12667,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2208-01	BU-703-COMP-06	BF142657.D	06 Jun 2025 20:32		RC/JU	Ok
20	Q2177-08	EB05312025	BF142658.D	06 Jun 2025 21:02		RC/JU	Ok
21	Q2242-01	TP09-MHJ	BF142659.D	06 Jun 2025 21:32		RC/JU	Ok
22	Q2202-03	MW-12-20250603	BF142660.D	06 Jun 2025 22:01	Need 10X Dilution	RC/JU	Dilution
23	Q2176-01	TP-46	BF142661.D	06 Jun 2025 22:31		RC/JU	Ok
24	Q2208-28	BU-703-COMP-09	BF142662.D	06 Jun 2025 23:01		RC/JU	Ok
25	DFTPP	DFTPP	BF142663.D	07 Jun 2025 00:32		RC/JU	Ok
26	SSTDCCC040	SSTDCCC040	BF142664.D	07 Jun 2025 01:38	CCC fail high for compounds #41,54,56,70,84 and failed low for compounds # 78,79,87,91,92.	RC/JU	Ok
27	PB168300BL	PB168300BL	BF142665.D	07 Jun 2025 02:08	CCC fail low for compound # 78,79,87,91,92.	RC/JU	Not Ok
28	PB168300BS	PB168300BS	BF142666.D	07 Jun 2025 02:38	CCC fail low for compound # 78,79,87,91,92. Also #92 Recovery fail low side.	RC/JU	Not Ok
29	Q2226-01	TP06-MHI-WC	BF142667.D	07 Jun 2025 03:08	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
30	Q2207-01	BU-703-COMP-01	BF142668.D	07 Jun 2025 03:38	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
31	Q2198-03	B-207-SB02	BF142669.D	07 Jun 2025 04:08	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
32	Q2176-03	TP-25	BF142670.D	07 Jun 2025 04:38	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
33	Q2176-05	TP-28	BF142671.D	07 Jun 2025 05:07	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
34	Q2125-07	GSB3	BF142672.D	07 Jun 2025 05:36	CCC fail low for compound #79	RC/JU	ReRun

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF060625

Review By	Rahul	Review On	6/9/2025 11:32:18 AM		
Supervise By	Jagrut	Supervise On	6/9/2025 12:07:45 PM		
SubDirectory	BF060625	HP Acquire Method	BNA_F	HP Processing Method	bf052025
STD. NAME		STD REF.#			
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791				
CCC	SP6787				
Internal Standard/PEM	S12667,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

35	Q2176-02	TP-56	BF142673.D	07 Jun 2025 06:05	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
36	Q2227-01	TP07-MHH-WC	BF142674.D	07 Jun 2025 06:35	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
37	Q2228-01	TP08-MHI-WC	BF142675.D	07 Jun 2025 07:03	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
38	Q2241-01	TP-N	BF142676.D	07 Jun 2025 07:32	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
39	Q2241-05	TP-S	BF142677.D	07 Jun 2025 08:02	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
40	Q2244-01	TP03-MHC	BF142678.D	07 Jun 2025 08:31	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
41	Q2244-01MS	TP03-MHCMS	BF142679.D	07 Jun 2025 09:00	CCC fail low for compound # 78,79,87,91,92.	RC/JU	Not Ok
42	Q2244-01MSD	TP03-MHCMSD	BF142680.D	07 Jun 2025 09:29	CCC fail low for compound # 78,79,87,91,92.	RC/JU	Not Ok
43	Q2223-01	HR-01-06042025	BF142681.D	07 Jun 2025 09:58	CCC fail low for compound # 78,79,87,91,92. Need Straight Run	RC/JU	ReRun
44	Q2223-03	HR-04-06042025	BF142682.D	07 Jun 2025 10:27	CCC fail low for compound # 78,79,87,91,92. Need Straight Run	RC/JU	ReRun
45	Q2224-01	EO-01-06042025	BF142683.D	07 Jun 2025 10:57	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
46	Q2225-01	SU-03-06042025	BF142684.D	07 Jun 2025 11:26	CCC fail low for compound # 78,79,87,91,92. Internal Standard Fail	RC/JU	ReRun
47	Q2206-01	TP-1	BF142685.D	07 Jun 2025 11:54	CCC fail low for compound # 78,79,87,91,92. Need Straight Run	RC/JU	ReRun
48	Q2195-01	OK-01-060325	BF142686.D	07 Jun 2025 12:23	CCC fail low for compound # 78,79,87,91,92. Internal Standard Fail	RC/JU	ReRun

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF060625

Review By	Rahul	Review On	6/9/2025 11:32:18 AM		
Supervise By	Jagrut	Supervise On	6/9/2025 12:07:45 PM		
SubDirectory	BF060625	HP Acquire Method	BNA_F	HP Processing Method	bf052025
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791				
CCC	SP6787				
Internal Standard/PEM	S12667,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF061125

Review By	Rahul	Review On	6/11/2025 9:54:45 AM
Supervise By	Jagrut	Supervise On	6/11/2025 9:58:22 AM
SubDirectory	BF061125	HP Acquire Method	BNA_F
		HP Processing Method	BF061125

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12668,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142710.D	10 Jun 2025 15:42		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142711.D	10 Jun 2025 16:11	A Fresh Calibration is required.	RC/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BF142712.D	10 Jun 2025 16:54		RC/JU	Ok
4	SSTDICC005	SSTDICC005	BF142713.D	10 Jun 2025 17:24		RC/JU	Ok
5	SSTDICC010	SSTDICC010	BF142714.D	10 Jun 2025 17:53		RC/JU	Ok,M
6	SSTDICC020	SSTDICC020	BF142715.D	10 Jun 2025 18:22		RC/JU	Ok
7	SSTDICCC040	SSTDICCC040	BF142716.D	10 Jun 2025 18:52	Calibration is Good for 8270 E, 8270 DOD and 625.1 methods.	RC/JU	Ok
8	SSTDICC050	SSTDICC050	BF142717.D	10 Jun 2025 19:21		RC/JU	Ok
9	SSTDICC060	SSTDICC060	BF142718.D	10 Jun 2025 19:50		RC/JU	Ok
10	SSTDICC080	SSTDICC080	BF142719.D	10 Jun 2025 20:19	Compound #09 removed from 80 PPM.	RC/JU	Ok,M
11	SSTDICV040	ICVBF061125	BF142720.D	10 Jun 2025 20:49		RC/JU	Ok
12	PB168323BL	PB168323BL	BF142721.D	10 Jun 2025 21:47		RC/JU	Ok
13	DFTPP	DFTPP	BF142722.D	11 Jun 2025 08:56		RC/JU	Ok
14	SSTDCCC040	SSTDCCC040	BF142723.D	11 Jun 2025 09:24		RC/JU	Ok
15	PB168376BL	PB168376BL	BF142724.D	11 Jun 2025 09:53		RC/JU	Ok
16	PB168376BS	PB168376BS	BF142725.D	11 Jun 2025 10:22		RC/JU	Ok,M
17	PB168234BS	PB168234BS	BF142726.D	11 Jun 2025 10:51		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF061125

Review By	Rahul	Review On	6/11/2025 9:54:45 AM		
Supervise By	Jagrut	Supervise On	6/11/2025 9:58:22 AM		
SubDirectory	BF061125	HP Acquire Method	BNA_F	HP Processing Method	BF061125
STD. NAME		STD REF.#			
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791				
CCC	SP6787				
Internal Standard/PEM	S12668,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	PB168285BS	PB168285BS	BF142727.D	11 Jun 2025 11:21		RC/JU	Ok,M
19	PB168285BSD	PB168285BSD	BF142728.D	11 Jun 2025 11:50		RC/JU	Ok,M
20	PB168378BS	PB168378BS	BF142729.D	11 Jun 2025 12:19		RC/JU	Ok,M
21	PB168378BSD	PB168378BSD	BF142730.D	11 Jun 2025 12:49		RC/JU	Ok,M
22	PB168378BL	PB168378BL	BF142731.D	11 Jun 2025 13:18		RC/JU	Ok
23	Q2264-04	EF-WW	BF142732.D	11 Jun 2025 13:52	Surrogate and Internal Standard Failed	RC/JU	Ok
24	Q2268-10	FB-20250605	BF142733.D	11 Jun 2025 14:21		RC/JU	Ok
25	Q2268-03	MW-2-20250605	BF142734.D	11 Jun 2025 14:50	Need 2X Dilution	RC/JU	Dilution
26	Q2268-04MS	MW-2-20250605MS	BF142735.D	11 Jun 2025 15:20		RC/JU	Ok,M
27	Q2268-05MSD	MW-2-20250605MSD	BF142736.D	11 Jun 2025 15:50		RC/JU	Ok
28	Q2268-06	MW-2-20250605-A	BF142737.D	11 Jun 2025 16:19	Need 2X Dilution	RC/JU	Dilution
29	Q2268-07	MW-6-20250605	BF142738.D	11 Jun 2025 16:49	Internal Standard Fail, Need 2X Dilution	RC/JU	Dilution
30	Q2268-08	MW-3-20250605	BF142739.D	11 Jun 2025 17:19	Internal Standard Fail, Need 2X Dilution	RC/JU	Dilution
31	Q2273-01	WC-4	BF142740.D	11 Jun 2025 17:49		RC/JU	Ok
32	Q2273-05	WC-6	BF142741.D	11 Jun 2025 18:18		RC/JU	Ok
33	Q2273-05MS	WC-6MS	BF142742.D	11 Jun 2025 18:48		RC/JU	Ok
34	Q2273-05MSD	WC-6MSD	BF142743.D	11 Jun 2025 19:18		RC/JU	Ok
35	Q2268-03DL	MW-2-20250605DL	BF142744.D	11 Jun 2025 19:48		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF061125

Review By	Rahul	Review On	6/11/2025 9:54:45 AM		
Supervise By	Jagrut	Supervise On	6/11/2025 9:58:22 AM		
SubDirectory	BF061125	HP Acquire Method	BNA_F	HP Processing Method	BF061125

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12668,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

36	Q2268-03	MW-2-20250605	BF142745.D	11 Jun 2025 20:17	Already analyzed with OK status	RC/JU	Not Ok
37	Q2280-01	VNJ-210	BF142746.D	11 Jun 2025 20:47	Internal standard fail	RC/JU	ReRun

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF061225

Review By	Rahul	Review On	6/13/2025 8:51:32 AM
Supervise By	mohammad	Supervise On	6/13/2025 9:04:03 AM
SubDirectory	BF061225	HP Acquire Method	BNA_F
		HP Processing Method	BF061125

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12668,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142747.D	12 Jun 2025 09:21		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142748.D	12 Jun 2025 09:50		RC/JU	Ok
3	PB168376BL	PB168376BL	BF142749.D	12 Jun 2025 10:56		RC/JU	Ok
4	PB168259BS	PB168259BS	BF142750.D	12 Jun 2025 11:25	Recovery fail for few compounds high side, need noncon	RC/JU	Ok
5	PB168412BL	PB168412BL	BF142751.D	12 Jun 2025 11:54		RC/JU	Ok
6	PB168412BS	PB168412BS	BF142752.D	12 Jun 2025 12:24	Recovery fail for few compounds high side, need noncon	RC/JU	Ok
7	Q2268-06DL	MW-2-20250605-ADL	BF142753.D	12 Jun 2025 13:00		RC/JU	Ok
8	Q2268-07DL	MW-6-20250605DL	BF142754.D	12 Jun 2025 13:29		RC/JU	Ok
9	Q2268-08DL	MW-3-20250605DL	BF142755.D	12 Jun 2025 13:59		RC/JU	Ok
10	Q2280-01	VNJ-210	BF142756.D	12 Jun 2025 14:29		RC/JU	Ok,M
11	Q2176-02	TP-56	BF142757.D	12 Jun 2025 14:58		RC/JU	Ok
12	Q2287-01	CONCRETE-SLAB	BF142758.D	12 Jun 2025 15:28		RC/JU	Ok
13	Q2296-09	WC-3	BF142759.D	12 Jun 2025 15:57		RC/JU	Ok
14	Q2296-09MS	WC-3MS	BF142760.D	12 Jun 2025 16:27		RC/JU	Ok
15	Q2296-09MSD	WC-3MSD	BF142761.D	12 Jun 2025 16:57	RPD fail for some compounds	RC/JU	Ok
16	Q2223-01	HR-01-06042025	BF142762.D	12 Jun 2025 17:27	Need noncon for as analyzed with 5x	RC/JU	Ok
17	Q2223-03	HR-04-06042025	BF142763.D	12 Jun 2025 17:57	Need noncon for as analyzed with 5x	RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF061225

Review By	Rahul	Review On	6/13/2025 8:51:32 AM		
Supervise By	mohammad	Supervise On	6/13/2025 9:04:03 AM		
SubDirectory	BF061225	HP Acquire Method	BNA_F	HP Processing Method	BF061125
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791				
CCC	SP6787				
Internal Standard/PEM	S12668,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2224-01	EO-01-06042025	BF142764.D	12 Jun 2025 18:27		RC/JU	Ok,M
19	Q2225-01	SU-03-06042025	BF142765.D	12 Jun 2025 18:57		RC/JU	Ok,M
20	Q2206-01	TP-1	BF142766.D	12 Jun 2025 19:26		RC/JU	Ok,M
21	Q2195-01	OK-01-060325	BF142767.D	12 Jun 2025 19:56		RC/JU	Ok,M
22	Q2283-01	MOO-25-0166-MOO-25	BF142768.D	12 Jun 2025 20:26		RC/JU	Ok,M
23	Q2278-01	TP-2	BF142769.D	12 Jun 2025 20:56	Need Straight Run	RC/JU	Not Ok
24	Q2286-01	LAW-25-0082-LAW-250	BF142770.D	12 Jun 2025 21:25	Out of Tune time.	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP060625

Review By	Rahul	Review On	6/9/2025 11:36:10 AM
Supervise By	Jagrut	Supervise On	6/9/2025 12:09:52 PM
SubDirectory	BP060625	HP Acquire Method	BNA_P
		HP Processing Method	BP060625

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12667,10ul/1000ul sample
ICV/I.BLK	SP6796
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024859.D	06 Jun 2025 09:49		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP024860.D	06 Jun 2025 10:30		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP024861.D	06 Jun 2025 11:11		RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP024862.D	06 Jun 2025 11:52		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP024863.D	06 Jun 2025 12:33	Calibration is Good for 8270 E, 8270 DOD and 625.1 methods.	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP024864.D	06 Jun 2025 13:14	Compound#54 & 56 are Kept on LR	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BP024865.D	06 Jun 2025 13:56		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BP024866.D	06 Jun 2025 14:37		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BP024867.D	06 Jun 2025 15:18		RC/JU	Ok
10	SSTDICV040	ICVBP060625	BP024868.D	06 Jun 2025 17:09		RC/JU	Ok,M
11	PB168259BL	PB168259BL	BP024869.D	06 Jun 2025 17:50		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP061025

Review By	anahy	Review On	6/11/2025 9:57:08 AM
Supervise By	Jagrut	Supervise On	6/11/2025 9:57:38 AM
SubDirectory	BP061025	HP Acquire Method	BNA_P
		HP Processing Method	BP060625

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12667,10ul/1000ul sample
ICV/I.BLK	SP6796
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024887.D	10 Jun 2025 10:12		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP024888.D	10 Jun 2025 10:53		RC/JU	Ok
3	PB168321BL	PB168321BL	BP024889.D	10 Jun 2025 11:33		RC/JU	Ok
4	PB168321BS	PB168321BS	BP024890.D	10 Jun 2025 12:14		RC/JU	Ok
5	Q2266-01	WC-3	BP024891.D	10 Jun 2025 13:02		RC/JU	Ok
6	Q2265-01	TP11-MHL-WC	BP024892.D	10 Jun 2025 13:43		RC/JU	Ok
7	Q2266-08RE	WC-5RE	BP024893.D	10 Jun 2025 14:24	Internal Standard Fail and Surrogate fail	RC/JU	Confirms
8	Q2176-08	TP-65	BP024894.D	10 Jun 2025 15:05		RC/JU	Ok
9	Q2177-02	B-187-SB01	BP024895.D	10 Jun 2025 15:46	Need 5X Dilution	RC/JU	Dilution
10	Q2161-02	B28-SOIL-SAMPLE	BP024896.D	10 Jun 2025 16:27	Surrogate Fail	RC/JU	Ok
11	Q2176-06	TP-27	BP024897.D	10 Jun 2025 17:08		RC/JU	Ok
12	Q2246-01	BU-03-060525	BP024898.D	10 Jun 2025 17:49		RC/JU	Ok,M
13	Q2177-06	B-202-SB01	BP024899.D	10 Jun 2025 18:30		RC/JU	Ok,M
14	Q2240-09	TP-1	BP024900.D	10 Jun 2025 19:10		RC/JU	Ok,M
15	Q2176-04	TP-26	BP024901.D	10 Jun 2025 19:51		RC/JU	Ok,M
16	Q2161-01	B27-SOIL-SAMPLE	BP024902.D	10 Jun 2025 20:32		RC/JU	Ok
17	Q2240-05	TP-2	BP024903.D	10 Jun 2025 21:13		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP061125

Review By	anahy	Review On	6/12/2025 9:31:34 AM
Supervise By	mohammad	Supervise On	6/13/2025 9:04:12 AM
SubDirectory	BP061125	HP Acquire Method	BNA_P
		HP Processing Method	BP060625

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12668,10ul/1000ul sample
ICV/I.BLK	SP6796
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024904.D	11 Jun 2025 09:29		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP024905.D	11 Jun 2025 10:09		RC/JU	Ok,M
3	PB168300BL	PB168300BL	BP024906.D	11 Jun 2025 10:50		RC/JU	Ok
4	PB168300BS	PB168300BS	BP024907.D	11 Jun 2025 11:31		RC/JU	Ok
5	Q2176-03	TP-25	BP024908.D	11 Jun 2025 12:18		RC/JU	Ok
6	Q2176-05	TP-28	BP024909.D	11 Jun 2025 12:59		RC/JU	Ok
7	Q2207-01	BU-703-COMP-01	BP024910.D	11 Jun 2025 13:40		RC/JU	Ok
8	Q2227-01	TP07-MHH-WC	BP024911.D	11 Jun 2025 14:21		RC/JU	Ok
9	Q2228-01	TP08-MHI-WC	BP024912.D	11 Jun 2025 15:02		RC/JU	Ok
10	Q2226-01	TP06-MHI-WC	BP024913.D	11 Jun 2025 15:43		RC/JU	Ok
11	Q2244-01	TP03-MHC	BP024914.D	11 Jun 2025 16:24		RC/JU	Ok
12	Q2244-01MS	TP03-MHCMS	BP024915.D	11 Jun 2025 17:05		RC/JU	Ok
13	Q2244-01MSD	TP03-MHCMSD	BP024916.D	11 Jun 2025 17:46		RC/JU	Ok
14	Q2177-02DL	B-187-SB01DL	BP024917.D	11 Jun 2025 18:27		RC/JU	Ok,M
15	Q2241-01	TP-N	BP024918.D	11 Jun 2025 19:08		RC/JU	Ok
16	Q2198-03	B-207-SB02	BP024919.D	11 Jun 2025 19:49		RC/JU	Ok,M
17	Q2125-07	GSB3	BP024920.D	11 Jun 2025 20:30		RC/JU	Ok
18	Q2241-05	TP-S	BP024921.D	11 Jun 2025 21:11		RC/JU	Ok

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP061125

Review By	anahy	Review On	6/12/2025 9:31:34 AM		
Supervise By	mohammad	Supervise On	6/13/2025 9:04:12 AM		
SubDirectory	BP061125	HP Acquire Method	BNA_P	HP Processing Method	BP060625
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791				
CCC	SP6787				
Internal Standard/PEM	S12668,10ul/1000ul sample				
ICV/I.BLK	SP6796				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

M : Manual Integration