

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/04/2025

Lab File ID: BF142602.D Time Analyzed: 12:27

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	118699	6.898	454762	8.18	238732	9.94
UPPER LIMIT	237398	7.398	909524	8.68	477464	10.439
LOWER LIMIT	59349.5	6.398	227381	7.68	119366	9.439
EPA SAMPLE NO.						
01 PB168234BL	131273	6.89	500088	8.18	273312	9.93
02 TP05-MHO-WCMS	119687	6.90	441391	8.18	218918	9.94
03 TP05-MHO-WCMSD	117119	6.90	429653	8.18	207287	9.94

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.