

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/11/2025

Lab File ID: BP024905.D Time Analyzed: 10:09

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	246136	7.607	989564	10.38	627529	14.25
UPPER LIMIT	492272	8.107	1979130	10.878	1255060	14.754
LOWER LIMIT	123068	7.107	494782	9.878	313765	13.754
EPA SAMPLE NO.						
01 TP-25	205856	7.61	780618	10.38	477183	14.25
02 TP-28	233414	7.61	929314	10.38	605225	14.25

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.