

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/10/2025

Lab File ID: BP024888.D Time Analyzed: 10:53

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1406880	17.048	1508870	21.466	1759120	24.707
UPPER LIMIT	2813760	17.548	3017740	21.966	3518240	25.207
LOWER LIMIT	703440	16.548	754435	20.966	879560	24.207
EPA SAMPLE NO.						
01 TP-27	1164370	17.06	1196930	21.49	1458420	24.74
02 TP-65	1373080	17.05	1382030	21.48	1592940	24.74
03 TP-26	1568910	17.06	1528630	21.50	1800540	24.77

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.