

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q2176

SDG NO.: Q2176

Lab File ID: BF142639.D

DFTPP Injection Date: 06/06/2025

Instrument ID: BNA\_F

DFTPP Injection Time: 11:01

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	37.0
68	Less than 2.0% of mass 69	0.5 ( 1.8 ) 1
69	Mass 69 relative abundance	32.3
70	Less than 2.0% of mass 69	0.2 ( 0.6 ) 1
127	10.0 - 80.0% of mass 198	45.2
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	28.0
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	17.0
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.9 ( 18.9 ) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142640.D	06/06/2025	11:36
TP-31	Q2176-07	BF142650.D	06/06/2025	17:01
TP-46	Q2176-01	BF142661.D	06/06/2025	22:31