

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q2176

SDG NO.: Q2176

Lab File ID: BP024904.D

DFTPP Injection Date: 06/11/2025

Instrument ID: BNA_P

DFTPP Injection Time: 09:29

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	29.9
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	34.6
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	45.8
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	30.8
365	Greater than 1% of mass 198	4.4
441	Present, but less than mass 443	13.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	16.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024905.D	06/11/2025	10:09
TP-25	Q2176-03	BP024908.D	06/11/2025	12:18
TP-28	Q2176-05	BP024909.D	06/11/2025	12:59