

Surrogate Summary

SDG No.: **Q2176**

Client: **CDM Smith**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2176-01	TP-46	1,2-Dichloroethane-d4	50	51.8	104	63	155
		Dibromofluoromethane	50	50.1	100	70	134
		Toluene-d8	50	49.5	99	74	123
		4-Bromofluorobenzene	50	43.1	86	17	146
Q2176-02	TP-56	1,2-Dichloroethane-d4	50	49.5	99	63	155
		Dibromofluoromethane	50	50.7	101	70	134
		Toluene-d8	50	49.2	98	74	123
		4-Bromofluorobenzene	50	40.3	81	17	146
Q2176-03	TP-25	1,2-Dichloroethane-d4	50	43.9	88	63	155
		Dibromofluoromethane	50	48.1	96	70	134
		Toluene-d8	50	49.4	99	74	123
		4-Bromofluorobenzene	50	39.2	78	17	146
Q2176-04	TP-26	1,2-Dichloroethane-d4	50	49.8	100	63	155
		Dibromofluoromethane	50	46.7	93	70	134
		Toluene-d8	50	49.7	99	74	123
		4-Bromofluorobenzene	50	43.1	86	17	146
Q2176-05	TP-28	1,2-Dichloroethane-d4	50	45.8	91	63	155
		Dibromofluoromethane	50	48.8	98	70	134
		Toluene-d8	50	49.9	100	74	123
		4-Bromofluorobenzene	50	40.8	82	17	146
Q2176-06	TP-27	1,2-Dichloroethane-d4	50	50.3	101	63	155
		Dibromofluoromethane	50	50.6	101	70	134
		Toluene-d8	50	49.8	100	74	123
		4-Bromofluorobenzene	50	42.1	84	17	146
Q2176-07	TP-31	1,2-Dichloroethane-d4	50	52.0	104	63	155
		Dibromofluoromethane	50	50.9	102	70	134
		Toluene-d8	50	49.7	99	74	123
		4-Bromofluorobenzene	50	41.4	83	17	146
Q2176-08	TP-65	1,2-Dichloroethane-d4	50	44.4	89	63	155
		Dibromofluoromethane	50	47.9	96	70	134
		Toluene-d8	50	49.6	99	74	123
		4-Bromofluorobenzene	50	38.2	76	17	146
VY0603SBL01	VY0603SBL01	1,2-Dichloroethane-d4	50	53.5	107	63	155
		Dibromofluoromethane	50	50.9	102	70	134
		Toluene-d8	50	49.9	100	74	123
		4-Bromofluorobenzene	50	42.0	84	17	146
VY0603SBS01	VY0603SBS01	1,2-Dichloroethane-d4	50	50.0	100	63	155
		Dibromofluoromethane	50	51.0	102	70	134
		Toluene-d8	50	51.5	103	74	123
		4-Bromofluorobenzene	50	49.0	98	17	146
VY0603SBSD01	VY0603SBSD01	1,2-Dichloroethane-d4	50	50.7	101	63	155
		Dibromofluoromethane	50	51.5	103	70	134
		Toluene-d8	50	52.6	105	74	123
		4-Bromofluorobenzene	50	50.1	100	17	146