

LB135986


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

38493



A= 0.0000e+000

B= 2.5552e-004

C= -1.8173e-002

Rho= 0.9994136

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	-0.028	-0.028	-40	0.000	-40	0				
0.2	0.200	0.213	0.013	903	0.0 %	903					
2.5	2.500	2.449	-0.051	9656	0.0 %	9656					
5.0	5.000	5.006	0.006	19663	0.0 %	19663					
7.5	7.500	7.743	0.243	30373	0.0 %	30373					
10.0	10.000	9.818	-0.182	38493	0.0 %	38493					

LB135986 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Jun 2025 09:26:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-40	-	0.0000	7470A	PPB	03 Jun 2025 10:09:52	S	Std
0.2 - 1	50.2	903	-	0.2000	7470A	PPB	03 Jun 2025 10:12:09	S	Std
2.5 - 1	52.5	9656	-	2.5000	7470A	PPB	03 Jun 2025 10:14:26	S	Std
5.0 - 1	55	19663	-	5.0000	7470A	PPB	03 Jun 2025 10:16:43	S	Std
7.5 - 1	57.5	30373	-	7.5000	7470A	PPB	03 Jun 2025 10:19:01	S	Std
10.0 - 1	50	38493	-	10.0000	7470A	PPB	03 Jun 2025 10:27:36	S	Std
ICV24 - 1	ICV24	16347	4.1589	-	7470A	PPB	03 Jun 2025 10:33:32	U	SMPL
ICB24 - 1	ICB24	-273	-0.0879	-	7470A	PPB	03 Jun 2025 10:35:47	U	SMPL
CCV86 - 1	CCV86	19065	4.8534	-	7470A	PPB	03 Jun 2025 10:38:04	U	SMPL
CCB86 - 1	CCB86	-176	-0.0631	-	7470A	PPB	03 Jun 2025 10:40:20	U	SMPL
CRA - 1	CRA	783	0.1819	-	7470A	PPB	03 Jun 2025 10:50:29	U	SMPL
HighStd - 1	HighStd	38704	9.8716	-	7470A	PPB	03 Jun 2025 10:52:44	U	SMPL
ChkStd - 1	ChkStd	24426	6.2232	-	7470A	PPB	03 Jun 2025 10:54:59	U	SMPL
PB168247BL - 1	PBW	179	0.0276	-	7470A	PPB	03 Jun 2025 10:59:59	U	SMPL
PB168247BS - 1	LCSW	13969	3.5512	-	7470A	PPB	03 Jun 2025 11:04:24	U	SMPL
Q2137-03 - 1	MOO-25-0149	568	0.1270	-	7470A	PPB	03 Jun 2025 11:06:39	U	SMPL
Q2155-03 - 1	447-WATER	1411	0.3424	-	7470A	PPB	03 Jun 2025 11:08:58	U	SMPL
Q2155-04 - 1	444	1193	0.2867	-	7470A	PPB	03 Jun 2025 11:11:15	U	SMPL
Q2155-04DUP - 1	444DUP	1008	0.2394	-	7470A	PPB	03 Jun 2025 11:13:33	U	SMPL
Q2155-04MS - 1	444MS	12754	3.2408	-	7470A	PPB	03 Jun 2025 11:20:39	U	SMPL
CCV87 - 1	CCV87	17929	4.5631	-	7470A	PPB	03 Jun 2025 11:22:54	U	SMPL
CCB87 - 1	CCB87	-298	-0.0943	-	7470A	PPB	03 Jun 2025 11:25:11	U	SMPL
Q2155-04MSD - 1	444MSD	13242	3.3655	-	7470A	PPB	03 Jun 2025 11:27:29	U	SMPL
Q2155-05 - 1	445	882	0.2072	-	7470A	PPB	03 Jun 2025 11:29:44	U	SMPL
Q2157-02 - 1	289	906	0.2133	-	7470A	PPB	03 Jun 2025 11:32:01	U	SMPL
Q2165-03 - 1	TRE-25-63-64	1079	0.2575	-	7470A	PPB	03 Jun 2025 11:34:17	U	SMPL
Q2165-06 - 1	TRE-25-65-66	2778	0.6917	-	7470A	PPB	03 Jun 2025 11:36:33	U	SMPL
Q2169-01 - 1	303-PPR-1	305013	77.9197	-	7470A	PPB	03 Jun 2025 11:38:49	U	SMPL
Q2169-03 - 1	303-PPR-2	8901	2.2562	-	7470A	PPB	03 Jun 2025 11:41:08	U	SMPL
Q2170-01 - 1	51925-AQ	1862	0.4576	-	7470A	PPB	03 Jun 2025 11:43:55	U	SMPL
Q2175-06 - 1	EGR-LIQUID	729	0.1681	-	7470A	PPB	03 Jun 2025 11:46:14	U	SMPL
Q2177-08 - 1	EB05312025	53	-0.0046	-	7470A	PPB	03 Jun 2025 11:48:32	U	SMPL
CCV88 - 1	CCV88	18876	4.8051	-	7470A	PPB	03 Jun 2025 11:50:49	U	SMPL
CCB88 - 1	CCB88	-145	-0.0552	-	7470A	PPB	03 Jun 2025 11:53:04	U	SMPL
Q2155-04LX5 - 1		163	0.0235	-	7470A	PPB	03 Jun 2025 11:55:22	U	SMPL
Q2155-04A - 1		12725	3.2334	-	7470A	PPB	03 Jun 2025 11:57:38	U	SMPL
Q2169-01DLX10 - 1		28120	7.1671	-	7470A	PPB	03 Jun 2025 11:59:53	U	SMPL
CCV89 - 1	CCV89	18871	4.8038	-	7470A	PPB	03 Jun 2025 12:02:10	U	SMPL
CCB89 - 1	CCB89	4	-0.0172	-	7470A	PPB	03 Jun 2025 12:04:29	U	SMPL