



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture **SDG No.:** Q2177
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2177 **SAS No.:** Q2177
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.18	0.2	91	70 - 130	CV	06/03/2025	10:50	LB135986
CRA	Mercury	0.21	0.2	104	70 - 130	CV	06/03/2025	12:52	LB135987
CRI01	Aluminum	98.3	100	98	65 - 135	P	06/04/2025	16:15	LB136011
	Antimony	50.4	50.0	101	65 - 135	P	06/04/2025	16:15	LB136011
	Arsenic	18.7	20.0	93	65 - 135	P	06/04/2025	16:15	LB136011
	Barium	86.6	100	87	65 - 135	P	06/04/2025	16:15	LB136011
	Beryllium	5.83	6.0	97	65 - 135	P	06/04/2025	16:15	LB136011
	Cadmium	5.99	6.0	100	65 - 135	P	06/04/2025	16:15	LB136011
	Calcium	1880	2000	94	65 - 135	P	06/04/2025	16:15	LB136011
	Chromium	10.8	10.0	108	65 - 135	P	06/04/2025	16:15	LB136011
	Cobalt	29.4	30.0	98	65 - 135	P	06/04/2025	16:15	LB136011
	Copper	22.0	20.0	110	65 - 135	P	06/04/2025	16:15	LB136011
	Iron	102	100	102	65 - 135	P	06/04/2025	16:15	LB136011
	Lead	11.2	12.0	93	65 - 135	P	06/04/2025	16:15	LB136011
	Magnesium	1930	2000	96	65 - 135	P	06/04/2025	16:15	LB136011
	Manganese	18.8	20.0	94	65 - 135	P	06/04/2025	16:15	LB136011
	Nickel	40.6	40.0	102	65 - 135	P	06/04/2025	16:15	LB136011
	Potassium	1840	2000	92	65 - 135	P	06/04/2025	16:15	LB136011
	Selenium	20.9	20.0	105	65 - 135	P	06/04/2025	16:15	LB136011
	Silver	10.0	10.0	100	65 - 135	P	06/04/2025	16:15	LB136011
	Sodium	1720	2000	86	65 - 135	P	06/04/2025	16:15	LB136011
	Thallium	38.5	40.0	96	65 - 135	P	06/04/2025	16:15	LB136011
	Vanadium	37.4	40.0	94	65 - 135	P	06/04/2025	16:15	LB136011
	Zinc	42.5	40.0	106	65 - 135	P	06/04/2025	16:15	LB136011
CRI01	Aluminum	105	100	105	65 - 135	P	06/10/2025	14:50	LB136097
	Antimony	53.3	50.0	107	65 - 135	P	06/10/2025	14:50	LB136097
	Arsenic	22.9	20.0	114	65 - 135	P	06/10/2025	14:50	LB136097
	Barium	98.7	100	99	65 - 135	P	06/10/2025	14:50	LB136097
	Beryllium	6.22	6.0	104	65 - 135	P	06/10/2025	14:50	LB136097
	Cadmium	6.07	6.0	101	65 - 135	P	06/10/2025	14:50	LB136097
	Calcium	2030	2000	102	65 - 135	P	06/10/2025	14:50	LB136097
	Chromium	10.7	10.0	107	65 - 135	P	06/10/2025	14:50	LB136097
	Cobalt	30.4	30.0	102	65 - 135	P	06/10/2025	14:50	LB136097
	Copper	22.7	20.0	114	65 - 135	P	06/10/2025	14:50	LB136097
	Iron	111	100	111	65 - 135	P	06/10/2025	14:50	LB136097

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture **SDG No.:** Q2177

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2177 **SAS No.:** Q2177

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Lead	13.4	12.0	112	65 - 135	P	06/10/2025	14:50	LB136097
	Magnesium	2040	2000	102	65 - 135	P	06/10/2025	14:50	LB136097
	Manganese	21.0	20.0	105	65 - 135	P	06/10/2025	14:50	LB136097
	Nickel	40.8	40.0	102	65 - 135	P	06/10/2025	14:50	LB136097
	Potassium	1990	2000	99	65 - 135	P	06/10/2025	14:50	LB136097
	Selenium	19.5	20.0	98	65 - 135	P	06/10/2025	14:50	LB136097
	Silver	10.6	10.0	106	65 - 135	P	06/10/2025	14:50	LB136097
	Sodium	1890	2000	95	65 - 135	P	06/10/2025	14:50	LB136097
	Thallium	40.8	40.0	102	65 - 135	P	06/10/2025	14:50	LB136097
	Vanadium	39.1	40.0	98	65 - 135	P	06/10/2025	14:50	LB136097
	Zinc	44.1	40.0	110	65 - 135	P	06/10/2025	14:50	LB136097