

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2177

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168247BL		WATER		Batch Number:	PB168247		Prep Date:	06/02/2025	
	Mercury	0.076	<0.2	U	0.20	CV	06/03/2025	10:59	LB135986
Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB168249BL		SOLID		Batch Number:	PB168249		Prep Date:	06/02/2025	
	Mercury	0.0070	<0.013	U	0.013	CV	06/03/2025	13:53	LB135987

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2177

Instrument: P4

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB168245BL	SOLID			Batch Number:	PB168245		Prep Date:	06/02/2025	
	Aluminum	0.84	<2.5	U	5.00	P	06/10/2025	20:35	LB136097
	Antimony	0.22	<1.25	U	2.50	P	06/10/2025	20:35	LB136097
	Arsenic	0.19	<0.5	U	1.00	P	06/10/2025	20:35	LB136097
	Barium	0.73	<2.5	U	5.00	P	06/10/2025	20:35	LB136097
	Beryllium	0.025	<0.15	U	0.30	P	06/10/2025	20:35	LB136097
	Cadmium	0.024	<0.15	U	0.30	P	06/10/2025	20:35	LB136097
	Calcium	11.1	<50	U	100	P	06/10/2025	20:35	LB136097
	Chromium	0.047	<0.25	U	0.50	P	06/10/2025	20:35	LB136097
	Cobalt	0.10	<0.75	U	1.50	P	06/10/2025	20:35	LB136097
	Copper	0.22	<0.5	U	1.00	P	06/10/2025	20:35	LB136097
	Iron	3.99	<2.5	U	5.00	P	06/10/2025	20:35	LB136097
	Lead	0.13	<0.3	U	0.60	P	06/10/2025	20:35	LB136097
	Magnesium	12.0	<50	U	100	P	06/10/2025	20:35	LB136097
	Manganese	0.14	<0.5	U	1.00	P	06/10/2025	20:35	LB136097
	Nickel	0.13	<1	U	2.00	P	06/10/2025	20:35	LB136097
	Potassium	27.7	<50	U	100	P	06/10/2025	20:35	LB136097
	Selenium	0.26	<0.5	U	1.00	P	06/10/2025	20:35	LB136097
	Silver	0.12	<0.25	U	0.50	P	06/10/2025	20:35	LB136097
	Sodium	17.8	<50	U	100	P	06/10/2025	20:35	LB136097
	Thallium	0.23	<1	U	2.00	P	06/10/2025	20:35	LB136097
	Vanadium	0.25	<1	U	2.00	P	06/10/2025	20:35	LB136097
	Zinc	0.23	<1	U	2.00	P	06/10/2025	20:35	LB136097
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168255BL	WATER			Batch Number:	PB168255		Prep Date:	06/03/2025	
	Aluminum	5.67	<25	U	50.0	P	06/04/2025	19:51	LB136011
	Antimony	3.38	<12.5	U	25.0	P	06/04/2025	19:51	LB136011
	Arsenic	2.56	<5	U	10.0	P	06/04/2025	19:51	LB136011
	Barium	7.28	<25	U	50.0	P	06/04/2025	19:51	LB136011
	Beryllium	0.28	<1.5	U	3.00	P	06/04/2025	19:51	LB136011
	Cadmium	0.25	<1.5	U	3.00	P	06/04/2025	19:51	LB136011
	Calcium	117	<500	U	1000	P	06/04/2025	19:51	LB136011
	Chromium	1.06	<2.5	U	5.00	P	06/04/2025	19:51	LB136011
	Cobalt	1.13	<7.5	U	15.0	P	06/04/2025	19:51	LB136011
	Copper	2.30	<5	U	10.0	P	06/04/2025	19:51	LB136011
	Iron	11.7	<25	U	50.0	P	06/04/2025	19:51	LB136011
	Lead	1.15	<3	U	6.00	P	06/04/2025	19:51	LB136011
	Magnesium	122	<500	U	1000	P	06/04/2025	19:51	LB136011
	Manganese	2.97	<5	U	10.0	P	06/04/2025	19:51	LB136011
	Nickel	1.53	<10	U	20.0	P	06/04/2025	19:51	LB136011

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2177

Instrument: P4

Potassium	459	<500	U	1000	P	06/04/2025	19:51	LB136011
Selenium	4.82	<5	U	10.0	P	06/04/2025	19:51	LB136011
Silver	0.81	<2.5	U	5.00	P	06/04/2025	19:51	LB136011
Sodium	434	<500	U	1000	P	06/04/2025	19:51	LB136011
Thallium	2.19	<10	U	20.0	P	06/04/2025	19:51	LB136011
Vanadium	3.13	<10	U	20.0	P	06/04/2025	19:51	LB136011
Zinc	8.33	<10	U	20.0	P	06/04/2025	19:51	LB136011