

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135995

Review By	MOHAN	Review On	6/4/2025 5:38:08 PM
Supervise By	jaswal	Supervise On	6/4/2025 5:41:58 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85853,MP85854,MP85855,MP85856,MP85857,MP85858		
ICV Standard	MP85859		
CCV Standard	MP85861		
ICSA Standard			
CRI Standard	MP85863		
LCS Standard			
Chk Standard	MP85860,MP85862,MP85864,MP85866		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/04/25 10:02		MOHAN	OK
2	S0.2	S0.2	CAL2	06/04/25 10:04		MOHAN	OK
3	S2.5	S2.5	CAL3	06/04/25 10:06		MOHAN	OK
4	S5	S5	CAL4	06/04/25 10:08		MOHAN	OK
5	S7.5	S7.5	CAL5	06/04/25 10:11		MOHAN	OK
6	S10	S10	CAL6	06/04/25 10:13		MOHAN	OK
7	ICV26	ICV26	ICV	06/04/25 10:25		MOHAN	OK
8	ICB26	ICB26	ICB	06/04/25 10:30		MOHAN	OK
9	CCV95	CCV95	CCV	06/04/25 10:32		MOHAN	OK
10	CCB95	CCB95	CCB	06/04/25 10:34		MOHAN	OK
11	CRA	CRA	CRDL	06/04/25 10:36		MOHAN	OK
12	HighStd	HighStd	HIGH STD	06/04/25 10:39		MOHAN	OK
13	ChkStd	ChkStd	SAM	06/04/25 10:41		MOHAN	OK
14	PB168262BL	PB168262BL	MB	06/04/25 10:43		MOHAN	OK
15	PB168262BS	PB168262BS	LCS	06/04/25 10:48		MOHAN	OK
16	Q2159-04	TP05-MHO-WC	SAM	06/04/25 10:50		MOHAN	OK
17	Q2160-04	TP04-MHG-WC	SAM	06/04/25 10:53		MOHAN	OK
18	Q2160-08	TP05-MHH-WC	SAM	06/04/25 10:55		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135995

Review By	MOHAN	Review On	6/4/2025 5:38:08 PM
Supervise By	jaswal	Supervise On	6/4/2025 5:41:58 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85853,MP85854,MP85855,MP85856,MP85857,MP85858		
ICV Standard	MP85859		
CCV Standard	MP85861		
ICSA Standard			
CRI Standard	MP85863		
LCS Standard			
Chk Standard	MP85860,MP85862,MP85864,MP85866		

19	Q2168-02	SAN-A1-A3	SAM	06/04/25 10:57		MOHAN	OK
20	Q2168-06	SAN-B1-B3	SAM	06/04/25 10:59		MOHAN	OK
21	CCV96	CCV96	CCV	06/04/25 11:02		MOHAN	OK
22	CCB96	CCB96	CCB	06/04/25 11:04		MOHAN	OK
23	Q2168-10	SAN-C1-C2	SAM	06/04/25 11:06		MOHAN	OK
24	Q2172-04	TP06-MHQ	SAM	06/04/25 11:09		MOHAN	OK
25	Q2173-06	OR-400-CF-402B-CO	SAM	06/04/25 11:11		MOHAN	OK
26	Q2173-12	OR-400-CF-402B-CO	SAM	06/04/25 11:13		MOHAN	OK
27	Q2173-18	OR-400-CF-402B-CO	SAM	06/04/25 11:15		MOHAN	OK
28	Q2177-03	B-187-SB01	SAM	06/04/25 11:18		MOHAN	OK
29	Q2177-05	B-187-SB02	SAM	06/04/25 11:20		MOHAN	OK
30	Q2177-07	B-202-SB01	SAM	06/04/25 11:22		MOHAN	OK
31	Q2185-04	TP02-MHB-WC	SAM	06/04/25 11:24		MOHAN	OK
32	Q2185-08	TP01-MHA-WC	SAM	06/04/25 11:27		MOHAN	OK
33	CCV97	CCV97	CCV	06/04/25 11:29		MOHAN	OK
34	CCB97	CCB97	CCB	06/04/25 11:31		MOHAN	OK
35	Q2185-08DUP	TP01-MHA-WC	DUP	06/04/25 11:34		MOHAN	OK
36	Q2185-08MS	TP01-MHA-WCMS	MS	06/04/25 11:38		MOHAN	OK
37	Q2185-08MSD	TP01-MHA-WCMSD	MSD	06/04/25 11:41		MOHAN	OK
38	PB168224TB	PB168224TB	MB	06/04/25 11:43		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135995

Review By	MOHAN	Review On	6/4/2025 5:38:08 PM
Supervise By	jaswal	Supervise On	6/4/2025 5:41:58 PM

STD. NAME	STD REF.#
ICAL Standard	MP85853,MP85854,MP85855,MP85856,MP85857,MP85858
ICV Standard	MP85859
CCV Standard	MP85861
ICSA Standard	
CRI Standard	MP85863
LCS Standard	
Chk Standard	MP85860,MP85862,MP85864,MP85866

39	Q2185-08L	TP01-MHA-WC	SD	06/04/25 11:45		MOHAN	OK
40	Q2185-08A	TP01-MHA-WC	PS	06/04/25 11:48		MOHAN	OK
41	CCV98	CCV98	CCV	06/04/25 11:53		MOHAN	OK
42	CCB98	CCB98	CCB	06/04/25 11:55		MOHAN	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136011

Review By	Janvi	Review On	6/5/2025 3:57:54 PM
Supervise By	jaswal	Supervise On	6/10/2025 4:55:25 PM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85872
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/04/25 14:18		Jaswal	OK
2	S1	S1	CAL2	06/04/25 14:23		Jaswal	OK
3	S2	S2	CAL3	06/04/25 14:27		Jaswal	OK
4	S3	S3	CAL4	06/04/25 14:31		Jaswal	OK
5	S4	S4	CAL5	06/04/25 14:35		Jaswal	OK
6	S5	S5	CAL6	06/04/25 14:39		Jaswal	OK
7	ICV01	ICV01	ICV	06/04/25 15:49		Jaswal	OK
8	LLICV01	LLICV01	LLICV	06/04/25 16:06		Jaswal	OK
9	ICB01	ICB01	ICB	06/04/25 16:10		Jaswal	OK
10	CRI01	CRI01	CRDL	06/04/25 16:15		Jaswal	OK
11	ICSA01	ICSA01	ICSA	06/04/25 16:26		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	06/04/25 16:32		Jaswal	OK
13	ICSADL	ICSADL	ICSA	06/04/25 16:36		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	06/04/25 16:40		Jaswal	OK
15	CCV01	CCV01	CCV	06/04/25 16:45		Jaswal	OK
16	CCB01	CCB01	CCB	06/04/25 16:52		Jaswal	OK
17	Q2160-05	TP05-MHH-WC	SAM	06/04/25 16:57		Jaswal	OK
18	Q2159-04	TP05-MHO-WC	SAM	06/04/25 17:01		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136011

Review By	Janvi	Review On	6/5/2025 3:57:54 PM
Supervise By	jaswal	Supervise On	6/10/2025 4:55:25 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85872		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

19	Q2178-01	RT2929	SAM	06/04/25 17:05		Jaswal	OK
20	Q2182-01	OR-03-06022025	SAM	06/04/25 17:10		Jaswal	OK
21	Q2172-04	TP06-MHQ	SAM	06/04/25 17:14		Jaswal	OK
22	Q2178-02	RT2929	SAM	06/04/25 17:18		Jaswal	OK
23	Q2185-01	TP02-MHB-WC	SAM	06/04/25 17:23		Jaswal	OK
24	Q2185-05	TP01-MHA-WC	SAM	06/04/25 17:27		Jaswal	OK
25	Q2178-01DUP	RT2929DUP	DUP	06/04/25 17:31		Jaswal	OK
26	CCV02	CCV02	CCV	06/04/25 17:35		Jaswal	OK
27	CCB02	CCB02	CCB	06/04/25 17:39		Jaswal	OK
28	Q2178-01L	RT2929L	SD	06/04/25 17:48		Jaswal	OK
29	Q2178-01MS	RT2929MS	MS	06/04/25 17:53		Jaswal	OK
30	Q2178-01MSD	RT2929MSD	MSD	06/04/25 17:57		Jaswal	OK
31	Q2178-01A	RT2929A	PS	06/04/25 18:01		Jaswal	OK
32	PB168224TB	PB168224TB	MB	06/04/25 18:05		Jaswal	OK
33	Q2168-02	SAN-A1-A3	SAM	06/04/25 18:10		Jaswal	OK
34	Q2168-06	SAN-B1-B3	SAM	06/04/25 18:14		Jaswal	OK
35	Q2168-10	SAN-C1-C2	SAM	06/04/25 18:19		Jaswal	OK
36	Q2177-03	B-187-SB01	SAM	06/04/25 18:23		Jaswal	OK
37	CCV03	CCV03	CCV	06/04/25 18:37		Jaswal	OK
38	CCB03	CCB03	CCB	06/04/25 18:42		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136011

Review By	Janvi	Review On	6/5/2025 3:57:54 PM
Supervise By	jaswal	Supervise On	6/10/2025 4:55:25 PM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85872
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

39	Q2177-05	B-187-SB02	SAM	06/04/25 18:47		Jaswal	OK
40	Q2177-07	B-202-SB01	SAM	06/04/25 18:51		Jaswal	OK
41	Q2185-04	TP02-MHB-WC	SAM	06/04/25 18:56		Jaswal	OK
42	Q2185-08	TP01-MHA-WC	SAM	06/04/25 19:00		Jaswal	OK
43	Q2185-08DUP	TP01-MHA-WC	DUP	06/04/25 19:04		Jaswal	OK
44	Q2185-08L	TP01-MHA-WC	SD	06/04/25 19:09		Jaswal	OK
45	Q2185-08MS	TP01-MHA-WCMS	MS	06/04/25 19:13		Jaswal	OK
46	Q2185-08MSD	TP01-MHA-WCMSD	MSD	06/04/25 19:17		Jaswal	OK
47	Q2185-08A	TP01-MHA-WC	PS	06/04/25 19:22		Jaswal	OK
48	CCV04	CCV04	CCV	06/04/25 19:26		Jaswal	OK
49	CCB04	CCB04	CCB	06/04/25 19:30		Jaswal	OK
50	PB168256BL	PB168256BL	MB	06/04/25 19:34		Jaswal	OK
51	PB168256BS	PB168256BS	LCS	06/04/25 19:39		Jaswal	OK
52	PB168261BL	PB168261BL	MB	06/04/25 19:43		Jaswal	OK
53	PB168261BS	PB168261BS	LCS	06/04/25 19:47		Jaswal	OK
54	PB168255BL	PB168255BL	MB	06/04/25 19:51		Jaswal	OK
55	PB168255BS	PB168255BS	LCS	06/04/25 19:55		Jaswal	OK
56	Q2169-01	303-PPR-1	SAM	06/04/25 19:59		Jaswal	OK
57	Q2169-03	303-PPR-2	SAM	06/04/25 20:03		Jaswal	OK
58	Q2175-06DL	EGR-LIQUIDDL	SAM	06/04/25 20:08	Straight 5x for all elements	Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136011

Review By	Janvi	Review On	6/5/2025 3:57:54 PM
Supervise By	jaswal	Supervise On	6/10/2025 4:55:25 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85872		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

59	CCV05	CCV05	CCV	06/04/25 20:12		Jaswal	OK
60	CCB05	CCB05	CCB	06/04/25 20:16		Jaswal	OK
61	Q2177-08	EB05312025	SAM	06/04/25 20:20		Jaswal	OK
62	Q2169-03DUP	303-PPR-2DUP	DUP	06/04/25 20:25		Jaswal	OK
63	Q2169-03L	303-PPR-2L	SD	06/04/25 20:29		Jaswal	OK
64	Q2169-03MS	303-PPR-2MS	MS	06/04/25 20:33		Jaswal	OK
65	Q2169-03MSD	303-PPR-2MSD	MSD	06/04/25 20:37		Jaswal	OK
66	Q2169-03A	303-PPR-2A	PS	06/04/25 20:42		Jaswal	OK
67	PB168276BL	PB168276BL	MB	06/04/25 20:46		Jaswal	OK
68	PB168276BS	PB168276BS	LCS	06/04/25 20:50		Jaswal	OK
69	Q2175-09DL	52725DL	SAM	06/04/25 20:54	Straight 5x for all elements	Jaswal	OK
70	CCV06	CCV06	CCV	06/04/25 20:58		Jaswal	OK
71	CCB06	CCB06	CCB	06/04/25 21:02		Jaswal	OK
72	Q2198-05DL	B-202-GW01DL	SAM	06/04/25 21:07	Straight 5x for all elements	Jaswal	OK
73	Q2198-05DUPDL	B-202-GW01DUPDL	DUP	06/04/25 21:11	Straight 5x for all elements	Jaswal	OK
74	Q2198-05LDL	B-202-GW01LDL	SD	06/04/25 21:15	Straight 25x for all elements	Jaswal	OK
75	Q2198-05MSDL	B-202-GW01MSDL	MS	06/04/25 21:20	Straight 5x for all elements	Jaswal	OK
76	Q2198-05MSDDL	B-202-GW01MSDDL	MSD	06/04/25 21:24	Straight 5x for all elements	Jaswal	OK
77	Q2198-05ADL	B-202-GW01ADL	PS	06/04/25 21:28	Straight 5x for all elements	Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QCBatch ID # LB136011

Review By	Janvi	Review On	6/5/2025 3:57:54 PM
Supervise By	jaswal	Supervise On	6/10/2025 4:55:25 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85872		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

78	Q2172-01	TP06-MHQ	SAM	06/04/25 21:32	CCV fail for Fe,K,Na	Jaswal	Not Ok
79	Q2172-01DUP	TP06-MHQDUP	DUP	06/04/25 21:37	CCV fail for Fe,K,Na	Jaswal	Not Ok
80	CCV07	CCV07	CCV	06/04/25 21:41		Jaswal	OK
81	CCB07	CCB07	CCB	06/04/25 21:45		Jaswal	OK
82	Q2172-01L	TP06-MHQL	SD	06/04/25 21:49	CCV fail for Fe,K,Na	Jaswal	Not Ok
83	Q2172-01MS	TP06-MHQMS	MS	06/04/25 21:54	CCV fail for Fe,K,Na	Jaswal	Not Ok
84	Q2172-01MSD	TP06-MHQMSD	MSD	06/04/25 21:57	CCV fail for Fe,K,Na	Jaswal	Not Ok
85	Q2172-01A	TP06-MHQA	PS	06/04/25 22:01	CCV fail for Fe,K,Na	Jaswal	Not Ok
86	CCV08	CCV08	CCV	06/04/25 22:06	CCV fail for Fe,K,Na	Jaswal	OK
87	CCB08	CCB08	CCB	06/04/25 22:10		Jaswal	OK