

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2177

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168224TB		WATER		Batch Number:	PB168262		Prep Date:	06/03/2025	
	Mercury	0.76	<2	U	2.00	CV	06/04/2025	11:43	LB135995
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168262BL		WATER		Batch Number:	PB168262		Prep Date:	06/03/2025	
	Mercury	0.076	<0.2	U	0.20	CV	06/04/2025	10:43	LB135995

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2177

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168224TB		WATER		Batch Number:	PB168256		Prep Date:	06/03/2025	
	Arsenic	25.6	<50	U	100	P	06/04/2025	18:05	LB136011
	Barium	72.8	<250	U	500	P	06/04/2025	18:05	LB136011
	Cadmium	2.50	<15	U	30.0	P	06/04/2025	18:05	LB136011
	Chromium	10.6	<25	U	50.0	P	06/04/2025	18:05	LB136011
	Lead	11.5	<30	U	60.0	P	06/04/2025	18:05	LB136011
	Selenium	48.2	<50	U	100	P	06/04/2025	18:05	LB136011
	Silver	8.10	<25	U	50.0	P	06/04/2025	18:05	LB136011
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168256BL		WATER		Batch Number:	PB168256		Prep Date:	06/03/2025	
	Arsenic	25.6	<50	U	100	P	06/04/2025	19:34	LB136011
	Barium	72.8	<250	U	500	P	06/04/2025	19:34	LB136011
	Cadmium	2.50	<15	U	30.0	P	06/04/2025	19:34	LB136011
	Chromium	10.6	<25	U	50.0	P	06/04/2025	19:34	LB136011
	Lead	11.5	<30	U	60.0	P	06/04/2025	19:34	LB136011
	Selenium	48.2	<50	U	100	P	06/04/2025	19:34	LB136011
	Silver	8.10	<25	U	50.0	P	06/04/2025	19:34	LB136011