

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Portal Partners Tri-Venture</u>	SDG No.: <u>Q2177</u>
Contract: <u>PORT06</u> Lab Code: <u>CHEM</u>	Case No.: <u>Q2177</u> SAS No.: <u>Q2177</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	1.52			-20	20	06/04/2025	16:26	LB136011
	Barium	0.52	6.0	9	-94	106	06/04/2025	16:26	LB136011
	Cadmium	0.92	1.0	92	-5	7	06/04/2025	16:26	LB136011
	Chromium	46.0	52.0	88	42	62	06/04/2025	16:26	LB136011
	Lead	-0.74			-12	12	06/04/2025	16:26	LB136011
	Selenium	1.80			-20	20	06/04/2025	16:26	LB136011
	Silver	4.01			-10	10	06/04/2025	16:26	LB136011
ICSAB01	Arsenic	106	104	102	88.4	120	06/04/2025	16:32	LB136011
	Barium	469	537	87	437	637	06/04/2025	16:32	LB136011
	Cadmium	1000	972	103	826	1120	06/04/2025	16:32	LB136011
	Chromium	556	542	103	460	624	06/04/2025	16:32	LB136011
	Lead	47.6	49.0	97	37	61	06/04/2025	16:32	LB136011
	Selenium	56.4	46.0	123	26	66	06/04/2025	16:32	LB136011
	Silver	219	201	109	170	232	06/04/2025	16:32	LB136011