


Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** Q2177
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2177 **SAS No.:** Q2177

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168256BS							
Arsenic	ug/L	4000	3770		94	80 - 120	P
Barium	ug/L	1000	882		88	80 - 120	P
Cadmium	ug/L	1000	949		95	80 - 120	P
Chromium	ug/L	2000	1910		96	80 - 120	P
Lead	ug/L	5000	4640		93	80 - 120	P
Selenium	ug/L	10000	9700		97	80 - 120	P
Silver	ug/L	380	342		90	80 - 120	P



- 7 -

Client:	Portal Partners Tri-Venture	SDG No.:	Q2177				
Contract:	PORT06	Lab Code:	CHEM	Case No.:	Q2177	SAS No.:	Q2177

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168262BS Mercury	ug/L	4.0	3.58		90	80 - 120	CV