

Continuing Calibration Report for SequenceID : FC061225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069183.D

Aliphatic C9-C12	6128822.000	60.000	3.305	6.606	102147.033	105717.491	3.377
Aliphatic C12-C16	3810297.000	40.000	6.607	10.010	95257.425	101985.435	6.597
Aliphatic C16-C21	5336149.000	60.000	10.011	13.380	88935.817	97865.081	9.124
Aliphatic C21-C28	7227579.000	80.000	13.381	17.048	90344.738	93803.915	3.688
Aliphatic C28-C40	10825792.000	120.000	17.049	22.036	90214.933	94439.035	4.473
Aliphatic EPH	33328639.000	360.000	3.305	22.036	92579.553	97587.136	5.131