

Continuing Calibration Report for SequenceID : FC061225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069191.D

Aliphatic C9-C12	5712750.000	60.000	3.305	6.606	95212.500	105717.491	9.937
Aliphatic C12-C16	3536714.000	40.000	6.607	10.010	88417.850	101985.435	13.303
Aliphatic C16-C21	4997704.000	60.000	10.011	13.380	83295.067	97865.081	14.888
Aliphatic C21-C28	7067909.000	80.000	13.381	17.048	88348.863	93803.915	5.815
Aliphatic C28-C40	11269620.000	120.000	17.049	22.036	93913.500	94439.035	0.556
Aliphatic EPH	32584697.000	360.000	3.305	22.036	90513.047	97587.136	7.249