

Continuing Calibration Report for SequenceID : FD061225AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049458.D

Aromatic C10-C12	6695899.000	40.000	4.424	6.168	167397.475	167344.833	-0.031
Aromatic C12-C16	10638865.000	60.000	6.169	8.797	177314.417	187525.524	5.445
Aromatic C16-C21	13638806.000	80.000	8.798	13.088	170485.075	184740.440	7.716
Aromatic C21-C36	27741823.000	180.000	13.089	18.510	154121.239	145057.816	-6.248
Aromatic EPH	58715393.000	360.000	4.424	18.510	163098.314	163430.463	0.203