

Continuing Calibration Report for SequenceID : FD061225AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049462.D

Aromatic C10-C12	6448171.000	40.000	4.424	6.168	161204.275	167344.833	3.669
Aromatic C12-C16	10337061.000	60.000	6.169	8.797	172284.350	187525.524	8.128
Aromatic C16-C21	13644805.000	80.000	8.798	13.088	170560.063	184740.440	7.676
Aromatic C21-C36	28370633.000	180.000	13.089	18.510	157614.628	145057.816	-8.656
Aromatic EPH	58800670.000	360.000	4.424	18.510	163335.194	163430.463	0.058