

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q2177 **SAS No :** Q2177 **SDG No:** Q2177
Sample No : PB168324BS **Datafile:** FC069135.D **Client ID :** PB168324BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aliphatic C9-C12	300	148	49		(40-140)
Aliphatic C12-C16	200	123	62		(40-140)
Aliphatic C16-C21	300	209	70		(40-140)
Aliphatic C21-C28	400	298	74		(40-140)
Aliphatic C28-C40	600	420	70		(40-140)