

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: PORT06

Lab Code: CHEM **Case No.:** Q2177 **SAS No.:** Q2177 **SDG NO.:** Q2177

Instrument ID: ECD_O **Date(s) Analyzed:** 05/14/2025 05/15/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.90	3.80	4.00	87106400
		2	3.98	3.88	4.08	63703800
		3	4.06	3.96	4.16	192327000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.06	3.96	4.16	153174000
		2	4.55	4.45	4.65	81946400
		3	4.79	4.69	4.89	152927000
		4	4.97	4.87	5.07	82251400
		5	5.01	4.91	5.11	54259200
Aroclor-1262	500	1	6.82	6.72	6.92	486664000
		2	7.33	7.23	7.43	808494000
		3	7.61	7.51	7.71	338264000
		4	7.67	7.57	7.77	590414000
		5	8.17	8.07	8.27	285216000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: PORT06

Lab Code: CHEM **Case No.:** Q2177 **SAS No.:** Q2177 **SDG NO.:** Q2177

Instrument ID: ECD_O **Date(s) Analyzed:** 05/14/2025 05/15/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.89	3.79	3.99	73843000
		2	3.98	3.88	4.08	54414200
		3	4.05	3.95	4.15	167101000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.05	3.95	4.15	130595000
		2	4.78	4.68	4.88	131160000
		3	4.95	4.85	5.05	68071600
		4	5.04	4.94	5.14	61676400
		5	5.21	5.11	5.31	65732600
Aroclor-1262	500	1	6.79	6.69	6.89	340780000
		2	7.29	7.19	7.39	502862000
		3	7.57	7.47	7.67	187635000
		4	7.64	7.54	7.74	330230000
		5	8.13	8.03	8.23	112306000