

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

Continuing Calib Date: 06/03/2025 Initial Calibration Date(s): 05/14/2025 05/15/2025

Continuing Calib Time: 22:13 Initial Calibration Time(s): 17:19 01:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-2	(2)	4.79	4.79	4.69	4.89	0.00
Aroclor-1016-3	(3)	4.84	4.85	4.75	4.95	0.01
Aroclor-1016-4	(4)	4.96	4.97	4.87	5.07	0.01
Aroclor-1016-5	(5)	5.22	5.23	5.13	5.33	0.01
Aroclor-1260-1	(1)	6.26	6.27	6.17	6.37	0.01
Aroclor-1260-2	(2)	6.45	6.46	6.36	6.56	0.01
Aroclor-1260-3	(3)	6.82	6.82	6.72	6.92	0.00
Aroclor-1260-4	(4)	7.08	7.08	6.98	7.18	0.01
Aroclor-1260-5	(5)	7.32	7.33	7.23	7.43	0.01
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.72	8.72	8.62	8.82	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

Continuing Calib Date: 06/03/2025 Initial Calibration Date(s): 05/14/2025 05/15/2025

Continuing Calib Time: 22:13 Initial Calibration Time(s): 17:19 01:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.75	4.76	4.66	4.86	0.01
Aroclor-1016-2	(2)	4.77	4.78	4.68	4.88	0.01
Aroclor-1016-3	(3)	4.95	4.95	4.85	5.05	0.00
Aroclor-1016-4	(4)	4.99	4.99	4.89	5.09	0.00
Aroclor-1016-5	(5)	5.20	5.21	5.11	5.31	0.01
Aroclor-1260-1	(1)	6.23	6.24	6.14	6.34	0.01
Aroclor-1260-2	(2)	6.42	6.43	6.33	6.53	0.01
Aroclor-1260-3	(3)	6.57	6.58	6.48	6.68	0.01
Aroclor-1260-4	(4)	7.04	7.05	6.95	7.15	0.01
Aroclor-1260-5	(5)	7.28	7.29	7.19	7.39	0.01
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.67	8.67	8.57	8.77	0.01