

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 06/03/2025

Lab Sample No.: AR1660CCC500 Data File : PO111433.D Time Analyzed: 16:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.768	4.673	4.873	552.970	500.000	10.6
Aroclor-1016-2	4.787	4.692	4.892	530.380	500.000	6.1
Aroclor-1016-3	4.844	4.749	4.949	530.020	500.000	6.0
Aroclor-1016-4	4.964	4.869	5.069	577.700	500.000	15.5
Aroclor-1016-5	5.221	5.126	5.326	582.000	500.000	16.4
Aroclor-1260-1	6.260	6.166	6.366	551.700	500.000	10.3
Aroclor-1260-2	6.450	6.355	6.555	560.720	500.000	12.1
Aroclor-1260-3	6.817	6.722	6.922	538.560	500.000	7.7
Aroclor-1260-4	7.076	6.983	7.183	507.550	500.000	1.5
Aroclor-1260-5	7.319	7.225	7.425	507.550	500.000	1.5
Decachlorobiphenyl	8.717	8.624	8.824	46.200	50.000	-7.6
Tetrachloro-m-xylene	3.680	3.582	3.782	52.150	50.000	4.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 06/03/2025

Lab Sample No.: AR1660CCC500 Data File : PO111433.D Time Analyzed: 16:42

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.754	4.658	4.858	489.210	500.000	-2.2
Aroclor-1016-2	4.773	4.677	4.877	498.300	500.000	-0.3
Aroclor-1016-3	4.948	4.852	5.052	493.440	500.000	-1.3
Aroclor-1016-4	4.990	4.894	5.094	474.950	500.000	-5.0
Aroclor-1016-5	5.203	5.107	5.307	496.560	500.000	-0.7
Aroclor-1260-1	6.233	6.138	6.338	477.340	500.000	-4.5
Aroclor-1260-2	6.420	6.326	6.526	472.940	500.000	-5.4
Aroclor-1260-3	6.573	6.479	6.679	441.750	500.000	-11.7
Aroclor-1260-4	7.044	6.949	7.149	457.770	500.000	-8.4
Aroclor-1260-5	7.285	7.191	7.391	448.540	500.000	-10.3
Decachlorobiphenyl	8.666	8.574	8.774	44.910	50.000	-10.2
Tetrachloro-m-xylene	3.676	3.579	3.779	51.120	50.000	2.2