

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL02 Date Analyzed: 06/03/2025

Lab Sample No.: AR1660CCC500 Data File : PO111447.D Time Analyzed: 22:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.767	4.673	4.873	506.430	500.000	1.3
Aroclor-1016-2	4.786	4.692	4.892	519.830	500.000	4.0
Aroclor-1016-3	4.843	4.749	4.949	548.120	500.000	9.6
Aroclor-1016-4	4.963	4.869	5.069	582.810	500.000	16.6
Aroclor-1016-5	5.220	5.126	5.326	594.000	500.000	18.8
Aroclor-1260-1	6.260	6.166	6.366	533.080	500.000	6.6
Aroclor-1260-2	6.449	6.355	6.555	563.270	500.000	12.7
Aroclor-1260-3	6.815	6.722	6.922	540.800	500.000	8.2
Aroclor-1260-4	7.075	6.983	7.183	513.040	500.000	2.6
Aroclor-1260-5	7.318	7.225	7.425	513.860	500.000	2.8
Decachlorobiphenyl	8.715	8.624	8.824	46.220	50.000	-7.6
Tetrachloro-m-xylene	3.680	3.582	3.782	51.410	50.000	2.8

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL02 Date Analyzed: 06/03/2025

Lab Sample No.: AR1660CCC500 Data File : PO111447.D Time Analyzed: 22:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.754	4.658	4.858	490.760	500.000	-1.8
Aroclor-1016-2	4.773	4.677	4.877	497.040	500.000	-0.6
Aroclor-1016-3	4.948	4.852	5.052	489.960	500.000	-2.0
Aroclor-1016-4	4.990	4.894	5.094	468.400	500.000	-6.3
Aroclor-1016-5	5.202	5.107	5.307	525.890	500.000	5.2
Aroclor-1260-1	6.232	6.138	6.338	477.990	500.000	-4.4
Aroclor-1260-2	6.420	6.326	6.526	472.480	500.000	-5.5
Aroclor-1260-3	6.572	6.479	6.679	441.840	500.000	-11.6
Aroclor-1260-4	7.043	6.949	7.149	451.090	500.000	-9.8
Aroclor-1260-5	7.284	7.191	7.391	436.650	500.000	-12.7
Decachlorobiphenyl	8.665	8.574	8.774	44.120	50.000	-11.8
Tetrachloro-m-xylene	3.676	3.579	3.779	51.100	50.000	2.2