

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 06/03/2025

Lab Sample No.: AR1660CCC500 Data File : PP072553.D Time Analyzed: 10:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.647	5.552	5.752	428.340	500.000	-14.3
Aroclor-1016-2	5.669	5.574	5.774	464.800	500.000	-7.0
Aroclor-1016-3	5.731	5.636	5.836	458.790	500.000	-8.2
Aroclor-1016-4	5.829	5.734	5.934	479.670	500.000	-4.1
Aroclor-1016-5	6.121	6.027	6.227	471.030	500.000	-5.8
Aroclor-1260-1	7.239	7.147	7.347	488.880	500.000	-2.2
Aroclor-1260-2	7.492	7.401	7.601	466.630	500.000	-6.7
Aroclor-1260-3	7.850	7.760	7.960	483.300	500.000	-3.3
Aroclor-1260-4	8.074	7.985	8.185	483.820	500.000	-3.2
Aroclor-1260-5	8.393	8.303	8.503	496.840	500.000	-0.6
Decachlorobiphenyl	10.195	10.113	10.313	50.460	50.000	0.9
Tetrachloro-m-xylene	4.495	4.396	4.596	48.810	50.000	-2.4

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 06/03/2025

Lab Sample No.: AR1660CCC500 Data File : PP072553.D Time Analyzed: 10:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.874	4.777	4.977	488.160	500.000	-2.4
Aroclor-1016-2	4.892	4.795	4.995	494.600	500.000	-1.1
Aroclor-1016-3	5.069	4.973	5.173	501.810	500.000	0.4
Aroclor-1016-4	5.111	5.015	5.215	504.360	500.000	0.9
Aroclor-1016-5	5.325	5.229	5.429	546.340	500.000	9.3
Aroclor-1260-1	6.357	6.264	6.464	520.360	500.000	4.1
Aroclor-1260-2	6.545	6.453	6.653	487.870	500.000	-2.4
Aroclor-1260-3	6.697	6.606	6.806	509.120	500.000	1.8
Aroclor-1260-4	7.168	7.077	7.277	511.170	500.000	2.2
Aroclor-1260-5	7.410	7.319	7.519	513.820	500.000	2.8
Decachlorobiphenyl	8.808	8.723	8.923	55.840	50.000	11.7
Tetrachloro-m-xylene	3.793	3.691	3.891	50.620	50.000	1.2