

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL04 Date Analyzed: 06/03/2025

Lab Sample No.: AR1660CCC500 Data File : PP072568.D Time Analyzed: 16:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.651	5.552	5.752	411.650	500.000	-17.7
Aroclor-1016-2	5.673	5.574	5.774	442.770	500.000	-11.4
Aroclor-1016-3	5.735	5.636	5.836	452.700	500.000	-9.5
Aroclor-1016-4	5.833	5.734	5.934	463.680	500.000	-7.3
Aroclor-1016-5	6.125	6.027	6.227	443.990	500.000	-11.2
Aroclor-1260-1	7.243	7.147	7.347	452.460	500.000	-9.5
Aroclor-1260-2	7.496	7.401	7.601	432.800	500.000	-13.4
Aroclor-1260-3	7.854	7.760	7.960	445.150	500.000	-11.0
Aroclor-1260-4	8.079	7.985	8.185	435.990	500.000	-12.8
Aroclor-1260-5	8.397	8.303	8.503	461.230	500.000	-7.8
Decachlorobiphenyl	10.199	10.113	10.313	47.320	50.000	-5.4
Tetrachloro-m-xylene	4.500	4.396	4.596	46.500	50.000	-7.0

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL04 Date Analyzed: 06/03/2025

Lab Sample No.: AR1660CCC500 Data File : PP072568.D Time Analyzed: 16:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.877	4.777	4.977	486.360	500.000	-2.7
Aroclor-1016-2	4.894	4.795	4.995	495.400	500.000	-0.9
Aroclor-1016-3	5.071	4.973	5.173	493.020	500.000	-1.4
Aroclor-1016-4	5.113	5.015	5.215	489.380	500.000	-2.1
Aroclor-1016-5	5.327	5.229	5.429	537.380	500.000	7.5
Aroclor-1260-1	6.358	6.264	6.464	503.150	500.000	0.6
Aroclor-1260-2	6.547	6.453	6.653	478.760	500.000	-4.2
Aroclor-1260-3	6.699	6.606	6.806	492.400	500.000	-1.5
Aroclor-1260-4	7.169	7.077	7.277	481.460	500.000	-3.7
Aroclor-1260-5	7.412	7.319	7.519	488.860	500.000	-2.2
Decachlorobiphenyl	8.810	8.723	8.923	53.130	50.000	6.3
Tetrachloro-m-xylene	3.795	3.691	3.891	50.510	50.000	1.0