

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL05 Date Analyzed: 06/03/2025

Lab Sample No.: AR1660CCC500 Data File : PP072583.D Time Analyzed: 22:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.650	5.552	5.752	414.970	500.000	-17.0
Aroclor-1016-2	5.672	5.574	5.774	445.260	500.000	-10.9
Aroclor-1016-3	5.734	5.636	5.836	446.030	500.000	-10.8
Aroclor-1016-4	5.832	5.734	5.934	459.390	500.000	-8.1
Aroclor-1016-5	6.124	6.027	6.227	450.950	500.000	-9.8
Aroclor-1260-1	7.242	7.147	7.347	467.980	500.000	-6.4
Aroclor-1260-2	7.495	7.401	7.601	442.440	500.000	-11.5
Aroclor-1260-3	7.853	7.760	7.960	445.300	500.000	-10.9
Aroclor-1260-4	8.076	7.985	8.185	438.500	500.000	-12.3
Aroclor-1260-5	8.395	8.303	8.503	458.750	500.000	-8.3
Decachlorobiphenyl	10.196	10.113	10.313	47.150	50.000	-5.7
Tetrachloro-m-xylene	4.499	4.396	4.596	46.390	50.000	-7.2

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL05 Date Analyzed: 06/03/2025

Lab Sample No.: AR1660CCC500 Data File : PP072583.D Time Analyzed: 22:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.874	4.777	4.977	491.730	500.000	-1.7
Aroclor-1016-2	4.892	4.795	4.995	501.620	500.000	0.3
Aroclor-1016-3	5.069	4.973	5.173	504.520	500.000	0.9
Aroclor-1016-4	5.111	5.015	5.215	495.960	500.000	-0.8
Aroclor-1016-5	5.325	5.229	5.429	570.870	500.000	14.2
Aroclor-1260-1	6.356	6.264	6.464	507.750	500.000	1.6
Aroclor-1260-2	6.545	6.453	6.653	474.610	500.000	-5.1
Aroclor-1260-3	6.697	6.606	6.806	489.600	500.000	-2.1
Aroclor-1260-4	7.167	7.077	7.277	507.640	500.000	1.5
Aroclor-1260-5	7.409	7.319	7.519	505.520	500.000	1.1
Decachlorobiphenyl	8.807	8.723	8.923	51.690	50.000	3.4
Tetrachloro-m-xylene	3.794	3.691	3.891	48.730	50.000	-2.5