

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL06 Date Analyzed: 06/04/2025

Lab Sample No.: AR1660CCC500 Data File : PP072598.D Time Analyzed: 03:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.651	5.552	5.752	412.490	500.000	-17.5
Aroclor-1016-2	5.673	5.574	5.774	443.380	500.000	-11.3
Aroclor-1016-3	5.735	5.636	5.836	442.870	500.000	-11.4
Aroclor-1016-4	5.832	5.734	5.934	457.390	500.000	-8.5
Aroclor-1016-5	6.124	6.027	6.227	458.290	500.000	-8.3
Aroclor-1260-1	7.241	7.147	7.347	486.770	500.000	-2.6
Aroclor-1260-2	7.495	7.401	7.601	454.090	500.000	-9.2
Aroclor-1260-3	7.853	7.760	7.960	452.580	500.000	-9.5
Aroclor-1260-4	8.077	7.985	8.185	450.360	500.000	-9.9
Aroclor-1260-5	8.395	8.303	8.503	469.550	500.000	-6.1
Decachlorobiphenyl	10.197	10.113	10.313	48.890	50.000	-2.2
Tetrachloro-m-xylene	4.499	4.396	4.596	46.450	50.000	-7.1

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL06 Date Analyzed: 06/04/2025

Lab Sample No.: AR1660CCC500 Data File : PP072598.D Time Analyzed: 03:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.875	4.777	4.977	489.340	500.000	-2.1
Aroclor-1016-2	4.893	4.795	4.995	496.050	500.000	-0.8
Aroclor-1016-3	5.069	4.973	5.173	496.260	500.000	-0.7
Aroclor-1016-4	5.112	5.015	5.215	481.770	500.000	-3.6
Aroclor-1016-5	5.325	5.229	5.429	585.340	500.000	17.1
Aroclor-1260-1	6.357	6.264	6.464	526.020	500.000	5.2
Aroclor-1260-2	6.546	6.453	6.653	483.260	500.000	-3.3
Aroclor-1260-3	6.698	6.606	6.806	502.650	500.000	0.5
Aroclor-1260-4	7.168	7.077	7.277	519.590	500.000	3.9
Aroclor-1260-5	7.410	7.319	7.519	517.000	500.000	3.4
Decachlorobiphenyl	8.808	8.723	8.923	53.670	50.000	7.3
Tetrachloro-m-xylene	3.794	3.691	3.891	49.930	50.000	-0.1