

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL07 Date Analyzed: 06/04/2025

Lab Sample No.: AR1660CCC500 Data File : PP072604.D Time Analyzed: 08:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.648	5.552	5.752	415.340	500.000	-16.9
Aroclor-1016-2	5.669	5.574	5.774	441.250	500.000	-11.8
Aroclor-1016-3	5.732	5.636	5.836	441.300	500.000	-11.7
Aroclor-1016-4	5.828	5.734	5.934	456.060	500.000	-8.8
Aroclor-1016-5	6.120	6.027	6.227	444.840	500.000	-11.0
Aroclor-1260-1	7.239	7.147	7.347	437.310	500.000	-12.5
Aroclor-1260-2	7.492	7.401	7.601	440.050	500.000	-12.0
Aroclor-1260-3	7.850	7.760	7.960	453.880	500.000	-9.2
Aroclor-1260-4	8.074	7.985	8.185	455.700	500.000	-8.9
Aroclor-1260-5	8.393	8.303	8.503	475.780	500.000	-4.8
Decachlorobiphenyl	10.193	10.113	10.313	48.420	50.000	-3.2
Tetrachloro-m-xylene	4.496	4.396	4.596	46.020	50.000	-8.0

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL07 Date Analyzed: 06/04/2025

Lab Sample No.: AR1660CCC500 Data File : PP072604.D Time Analyzed: 08:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.874	4.777	4.977	467.590	500.000	-6.5
Aroclor-1016-2	4.893	4.795	4.995	476.230	500.000	-4.8
Aroclor-1016-3	5.069	4.973	5.173	481.410	500.000	-3.7
Aroclor-1016-4	5.111	5.015	5.215	490.810	500.000	-1.8
Aroclor-1016-5	5.324	5.229	5.429	482.150	500.000	-3.6
Aroclor-1260-1	6.356	6.264	6.464	503.880	500.000	0.8
Aroclor-1260-2	6.544	6.453	6.653	481.760	500.000	-3.6
Aroclor-1260-3	6.696	6.606	6.806	518.660	500.000	3.7
Aroclor-1260-4	7.167	7.077	7.277	515.270	500.000	3.1
Aroclor-1260-5	7.409	7.319	7.519	519.290	500.000	3.9
Decachlorobiphenyl	8.807	8.723	8.923	53.590	50.000	7.2
Tetrachloro-m-xylene	3.794	3.691	3.891	48.580	50.000	-2.8