

Surrogate Summary

SDG No.: Q2177

Client: Portal Partners Tri-Venture

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO111057.D	PIBLK-PO111057.D	Tetrachloro-m-xylene	1	20	16.4	82		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	17.5	87		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	15.4	77		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	17.9	90		70 (60)	130 (140)
I.BLK-PO111437.D	PIBLK-PO111437.D	Tetrachloro-m-xylene	1	20	20.1	100		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	19.4	97		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	18.6	93		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	18.2	91		70 (60)	130 (140)
Q2177-08	EB05312025	Tetrachloro-m-xylene	1	20	22.8	114		30 (30)	150 (173)
		Decachlorobiphenyl	1	20	15.4	77		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	21.0	105		30 (30)	150 (173)
		Decachlorobiphenyl	2	20	15.0	75		30 (10)	150 (173)
I.BLK-PO111451.D	PIBLK-PO111451.D	Tetrachloro-m-xylene	1	20	19.8	99		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	19.0	95		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	18.3	92		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	17.8	89		70 (60)	130 (140)
I.BLK-PP072163.D	PIBLK-PP072163.D	Tetrachloro-m-xylene	1	20	16.3	82		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	16.7	84		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	16.6	83		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	17.0	85		70 (60)	130 (140)
I.BLK-PP072557.D	PIBLK-PP072557.D	Tetrachloro-m-xylene	1	20	19.2	96		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	20.2	101		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.9	104		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	23.4	117		70 (60)	130 (140)
PB168252BL	PB168252BL	Tetrachloro-m-xylene	1	20	19.4	97		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.5	102		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.1	105		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	22.6	113		30 (32)	150 (175)
I.BLK-PP072572.D	PIBLK-PP072572.D	Tetrachloro-m-xylene	1	20	18.9	95		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	19.7	99		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.3	102		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	22.7	114		70 (60)	130 (140)
Q2176-05MS	TP-28MS	Tetrachloro-m-xylene	1	20	21.4	107		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.3	97		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	22.3	112		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	22.2	111		30 (32)	150 (175)
Q2176-05MSD	TP-28MSD	Tetrachloro-m-xylene	1	20	18.8	94		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	16.1	80		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.6	98		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	19.3	97		30 (32)	150 (175)
Q2177-02	B-187-SB01	Tetrachloro-m-xylene	1	20	18.4	92		30 (32)	150 (144)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q2177

Client: Portal Partners Tri-Venture

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q2177-02	B-187-SB01	Decachlorobiphenyl	1	20	14.6	73		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.7	104		30 (32)	150 (144)
Q2177-04	B-187-SB02	Decachlorobiphenyl	2	20	17.4	87		30 (32)	150 (175)
		Tetrachloro-m-xylene	1	20	18.0	90		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	13.4	67		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.8	104		30 (32)	150 (144)
Q2177-06	B-202-SB01	Decachlorobiphenyl	2	20	15.4	77		30 (32)	150 (175)
		Tetrachloro-m-xylene	1	20	20.5	102		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	18.7	94		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	24.1	120		30 (32)	150 (144)
I.BLK-PP072587.D	PIBLK-PP072587.D	Decachlorobiphenyl	2	20	21.8	109		30 (32)	150 (175)
		Tetrachloro-m-xylene	1	20	18.6	93		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	19.5	98		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.5	103		70 (60)	130 (140)
PB168257BL	PB168257BL	Decachlorobiphenyl	2	20	23.0	115		70 (60)	130 (140)
		Tetrachloro-m-xylene	1	20	19.0	95		30 (30)	150 (173)
		Decachlorobiphenyl	1	20	20.9	104		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	20.5	103		30 (30)	150 (173)
I.BLK-PP072602.D	PIBLK-PP072602.D	Decachlorobiphenyl	2	20	24.1	121		30 (10)	150 (173)
		Tetrachloro-m-xylene	1	20	18.5	93		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	20.3	101		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.5	102		70 (60)	130 (140)
I.BLK-PP072608.D	PIBLK-PP072608.D	Decachlorobiphenyl	2	20	22.8	114		70 (60)	130 (140)
		Tetrachloro-m-xylene	1	20	16.9	85		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	18.2	91		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	19.0	95		70 (60)	130 (140)
PB168257BS	PB168257BS	Decachlorobiphenyl	2	20	21.9	109		70 (60)	130 (140)
		Tetrachloro-m-xylene	1	20	19.4	97		30 (30)	150 (173)
		Decachlorobiphenyl	1	20	19.8	99		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	20.0	100		30 (30)	150 (173)
PB168257BSD	PB168257BSD	Decachlorobiphenyl	2	20	23.9	119		30 (10)	150 (173)
		Tetrachloro-m-xylene	1	20	19.3	96		30 (30)	150 (173)
		Decachlorobiphenyl	1	20	20.8	104		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	19.8	99		30 (30)	150 (173)
PB168252BS	PB168252BS	Decachlorobiphenyl	2	20	23.5	117		30 (10)	150 (173)
		Tetrachloro-m-xylene	1	20	19.5	97		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.5	103		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.0	100		30 (32)	150 (144)
I.BLK-PP072623.D	PIBLK-PP072623.D	Decachlorobiphenyl	2	20	23.5	117		30 (32)	150 (175)
		Tetrachloro-m-xylene	1	20	17.3	87		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	18.9	94		70 (60)	130 (140)

Surrogate Summary

SDG No.: Q2177

Client: Portal Partners Tri-Venture

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP072623.D	PIBLK-PP072623.D	Tetrachloro-m-xylene	2	20	19.3	97		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	21.0	105		70 (60)	130 (140)