

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL01 Date Analyzed: 06/04/2025

Lab Sample No.: HSTDCCC750 Data File : PS030497.D Time Analyzed: 20:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.398	9.304	9.504	724.280	712.500	1.7
2,4-D	8.506	8.412	8.612	705.370	705.000	0.1
2,4-DCAA	7.365	7.269	7.469	725.600	750.000	-3.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL01 Date Analyzed: 06/04/2025

Lab Sample No.: HSTDCCC750 Data File : PS030497.D Time Analyzed: 20:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.939	9.842	10.042	709.400	712.500	-0.4
2,4-D	9.032	8.934	9.134	695.700	705.000	-1.3
2,4-DCAA	7.772	7.674	7.874	722.520	750.000	-3.7