

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL02 Date Analyzed: 06/05/2025

Lab Sample No.: HSTDCCC750 Data File : PS030506.D Time Analyzed: 00:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.396	9.304	9.504	710.030	712.500	-0.3
2,4-D	8.504	8.412	8.612	681.590	705.000	-3.3
2,4-DCAA	7.364	7.269	7.469	717.850	750.000	-4.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL02 Date Analyzed: 06/05/2025

Lab Sample No.: HSTDCCC750 Data File : PS030506.D Time Analyzed: 00:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.939	9.842	10.042	704.930	712.500	-1.1
2,4-D	9.032	8.934	9.134	692.970	705.000	-1.7
2,4-DCAA	7.772	7.674	7.874	719.760	750.000	-4.0