

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL03 Date Analyzed: 06/05/2025

Lab Sample No.: HSTDCCC750 Data File : PS030516.D Time Analyzed: 04:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.395	9.304	9.504	720.680	712.500	1.1
2,4-D	8.504	8.412	8.612	697.490	705.000	-1.1
2,4-DCAA	7.364	7.269	7.469	713.260	750.000	-4.9

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL03 Date Analyzed: 06/05/2025

Lab Sample No.: HSTDCCC750 Data File : PS030516.D Time Analyzed: 04:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.939	9.842	10.042	709.690	712.500	-0.4
2,4-D	9.031	8.934	9.134	727.630	705.000	3.2
2,4-DCAA	7.772	7.674	7.874	709.210	750.000	-5.4