

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

Continuing Calib Date: 06/03/2025 Initial Calibration Date(s): 05/21/2025 05/21/2025

Continuing Calib Time: 20:27 Initial Calibration Time(s): 11:35 12:29

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.10	9.10	9.00	9.20	0.00
Tetrachloro-m-xylene	3.57	3.57	3.47	3.67	0.00
gamma-BHC (Lindane)	4.36	4.36	4.26	4.46	0.00
Heptachlor	4.95	4.96	4.86	5.06	0.01
Heptachlor epoxide	5.72	5.72	5.62	5.82	0.00
Endrin	6.60	6.60	6.50	6.70	0.00
Methoxychlor	7.52	7.52	7.42	7.62	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

Continuing Calib Date: 06/03/2025 Initial Calibration Date(s): 05/21/2025 05/21/2025

Continuing Calib Time: 20:27 Initial Calibration Time(s): 11:35 12:29

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.89	2.89	2.79	2.99	0.01
gamma-BHC (Lindane)	3.73	3.73	3.63	3.83	0.00
Heptachlor	4.08	4.08	3.98	4.18	0.00
Heptachlor epoxide	4.87	4.87	4.77	4.97	0.00
Endrin	5.78	5.78	5.68	5.88	0.00
Methoxychlor	6.75	6.75	6.65	6.85	0.00