

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

Continuing Calib Date: 06/04/2025 Initial Calibration Date(s): 05/21/2025 05/21/2025

Continuing Calib Time: 11:26 Initial Calibration Time(s): 11:35 12:29

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Decachlorobiphenyl   | 9.10       | 9.10      | 9.00      | 9.20 | 0.00       |
| Tetrachloro-m-xylene | 3.57       | 3.57      | 3.47      | 3.67 | 0.00       |
| gamma-BHC (Lindane)  | 4.36       | 4.36      | 4.26      | 4.46 | 0.00       |
| Heptachlor           | 4.95       | 4.96      | 4.86      | 5.06 | 0.01       |
| Heptachlor epoxide   | 5.72       | 5.72      | 5.62      | 5.82 | 0.00       |
| Endrin               | 6.60       | 6.60      | 6.50      | 6.70 | 0.00       |
| Methoxychlor         | 7.52       | 7.52      | 7.42      | 7.62 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

Continuing Calib Date: 06/04/2025 Initial Calibration Date(s): 05/21/2025 05/21/2025

Continuing Calib Time: 11:26 Initial Calibration Time(s): 11:35 12:29

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Decachlorobiphenyl   | 8.06       | 8.06      | 7.96                   | 8.16 | 0.00       |
| Tetrachloro-m-xylene | 2.89       | 2.89      | 2.79                   | 2.99 | 0.00       |
| gamma-BHC (Lindane)  | 3.73       | 3.73      | 3.63                   | 3.83 | 0.00       |
| Heptachlor           | 4.08       | 4.08      | 3.98                   | 4.18 | 0.00       |
| Heptachlor epoxide   | 4.87       | 4.87      | 4.77                   | 4.97 | 0.00       |
| Endrin               | 5.78       | 5.78      | 5.68                   | 5.88 | 0.00       |
| Methoxychlor         | 6.75       | 6.75      | 6.65                   | 6.85 | 0.00       |