

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL01 Date Analyzed: 06/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL095897.D Time Analyzed: 17:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.098	9.003	9.203	43.980	50.000	-12.0
Endrin	6.601	6.502	6.702	44.600	50.000	-10.8
gamma-BHC (Lindane)	4.356	4.257	4.457	49.830	50.000	-0.3
Heptachlor	4.953	4.855	5.055	48.950	50.000	-2.1
Heptachlor epoxide	5.717	5.618	5.818	48.650	50.000	-2.7
Methoxychlor	7.520	7.423	7.623	44.540	50.000	-10.9
Tetrachloro-m-xylene	3.572	3.473	3.673	50.900	50.000	1.8

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL01 Date Analyzed: 06/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL095897.D Time Analyzed: 17:01

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.059	7.964	8.164	42.370	50.000	-15.3
Endrin	5.779	5.683	5.883	45.120	50.000	-9.8
gamma-BHC (Lindane)	3.729	3.630	3.830	52.270	50.000	4.5
Heptachlor	4.081	3.983	4.183	51.390	50.000	2.8
Heptachlor epoxide	4.868	4.771	4.971	49.980	50.000	0.0
Methoxychlor	6.747	6.651	6.851	40.990	50.000	-18.0
Tetrachloro-m-xylene	2.886	2.786	2.986	51.450	50.000	2.9