

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL02 Date Analyzed: 06/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL095907.D Time Analyzed: 19:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.097	9.003	9.203	49.020	50.000	-2.0
Endrin	6.599	6.502	6.702	48.510	50.000	-3.0
gamma-BHC (Lindane)	4.355	4.257	4.457	55.210	50.000	10.4
Heptachlor	4.953	4.855	5.055	53.750	50.000	7.5
Heptachlor epoxide	5.716	5.618	5.818	54.140	50.000	8.3
Methoxychlor	7.519	7.423	7.623	48.380	50.000	-3.2
Tetrachloro-m-xylene	3.572	3.473	3.673	54.960	50.000	9.9

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL02 Date Analyzed: 06/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL095907.D Time Analyzed: 19:18

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.059	7.964	8.164	51.620	50.000	3.2
Endrin	5.778	5.683	5.883	51.740	50.000	3.5
gamma-BHC (Lindane)	3.729	3.630	3.830	58.190	50.000	16.4
Heptachlor	4.081	3.983	4.183	56.080	50.000	12.2
Heptachlor epoxide	4.867	4.771	4.971	57.480	50.000	15.0
Methoxychlor	6.746	6.651	6.851	47.170	50.000	-5.7
Tetrachloro-m-xylene	2.885	2.786	2.986	57.170	50.000	14.3