

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL03 Date Analyzed: 06/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL095912.D Time Analyzed: 20:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.098	9.003	9.203	49.580	50.000	-0.8
Endrin	6.599	6.502	6.702	49.670	50.000	-0.7
gamma-BHC (Lindane)	4.355	4.257	4.457	55.690	50.000	11.4
Heptachlor	4.953	4.855	5.055	54.510	50.000	9.0
Heptachlor epoxide	5.715	5.618	5.818	53.980	50.000	8.0
Methoxychlor	7.519	7.423	7.623	49.420	50.000	-1.2
Tetrachloro-m-xylene	3.572	3.473	3.673	55.240	50.000	10.5

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL03 Date Analyzed: 06/03/2025

Lab Sample No.: PSTDCCC050 Data File : PL095912.D Time Analyzed: 20:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.059	7.964	8.164	53.360	50.000	6.7
Endrin	5.778	5.683	5.883	51.550	50.000	3.1
gamma-BHC (Lindane)	3.728	3.630	3.830	58.370	50.000	16.7
Heptachlor	4.081	3.983	4.183	56.590	50.000	13.2
Heptachlor epoxide	4.867	4.771	4.971	57.880	50.000	15.8
Methoxychlor	6.746	6.651	6.851	48.250	50.000	-3.5
Tetrachloro-m-xylene	2.885	2.786	2.986	57.160	50.000	14.3