

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL04 Date Analyzed: 06/04/2025

Lab Sample No.: PSTDCCC050 Data File : PL095916.D Time Analyzed: 11:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.098	9.003	9.203	50.250	50.000	0.5
Endrin	6.600	6.502	6.702	48.710	50.000	-2.6
gamma-BHC (Lindane)	4.355	4.257	4.457	56.730	50.000	13.5
Heptachlor	4.953	4.855	5.055	55.030	50.000	10.1
Heptachlor epoxide	5.716	5.618	5.818	55.620	50.000	11.2
Methoxychlor	7.520	7.423	7.623	49.270	50.000	-1.5
Tetrachloro-m-xylene	3.572	3.473	3.673	56.700	50.000	13.4

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL04 Date Analyzed: 06/04/2025

Lab Sample No.: PSTDCCC050 Data File : PL095916.D Time Analyzed: 11:26

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.060	7.964	8.164	52.030	50.000	4.1
Endrin	5.780	5.683	5.883	48.760	50.000	-2.5
gamma-BHC (Lindane)	3.728	3.630	3.830	57.100	50.000	14.2
Heptachlor	4.081	3.983	4.183	55.210	50.000	10.4
Heptachlor epoxide	4.868	4.771	4.971	55.730	50.000	11.5
Methoxychlor	6.747	6.651	6.851	46.180	50.000	-7.6
Tetrachloro-m-xylene	2.886	2.786	2.986	56.010	50.000	12.0