

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL05 Date Analyzed: 06/04/2025

Lab Sample No.: PSTDCCC050 Data File : PL095919.D Time Analyzed: 13:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.099	9.003	9.203	50.390	50.000	0.8
Endrin	6.600	6.502	6.702	48.640	50.000	-2.7
gamma-BHC (Lindane)	4.355	4.257	4.457	56.510	50.000	13.0
Heptachlor	4.952	4.855	5.055	55.130	50.000	10.3
Heptachlor epoxide	5.716	5.618	5.818	55.620	50.000	11.2
Methoxychlor	7.520	7.423	7.623	49.710	50.000	-0.6
Tetrachloro-m-xylene	3.571	3.473	3.673	56.360	50.000	12.7

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL05 Date Analyzed: 06/04/2025

Lab Sample No.: PSTDCCC050 Data File : PL095919.D Time Analyzed: 13:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.058	7.964	8.164	52.910	50.000	5.8
Endrin	5.778	5.683	5.883	49.690	50.000	-0.6
gamma-BHC (Lindane)	3.727	3.630	3.830	57.470	50.000	14.9
Heptachlor	4.079	3.983	4.183	55.520	50.000	11.0
Heptachlor epoxide	4.866	4.771	4.971	56.660	50.000	13.3
Methoxychlor	6.746	6.651	6.851	47.180	50.000	-5.6
Tetrachloro-m-xylene	2.884	2.786	2.986	56.310	50.000	12.6