

Report of Analysis

Client:	Portal Partners Tr	i-Venture			Date Collected:	05/31/25		
Project:	Amtrak Sawtooth	ooth Bridges 2025			Date Received:	06/02/25		
Client Sample ID:	B-187-SB02				SDG No.:	Q2177		
Lab Sample ID:	Q2177-05				Matrix:	TCLP		
Analytical Method					% Solid:	0	Decanted:	
-								
Sample Wt/Vol:	100 Units:	mL			Final Vol:	10000	uL	
Soil Aliquot Vol:		uL			Test:	TCLP Pestic	ide	
Extraction Type:					Injection Volume :			
GPC Factor :	1.0	PH :						
Prep Method :	SW3541B							
Trep Method .	0110							
File ID/Qc Batch:	Dilution:	Prep	Prep Date		Date Analyzed	Prep Batch ID		
PL095909.D	1	06/03/25 11:34			06/03/25 19:46	PB168264		
CAS Number	Parameter	Conc.	Qualifier	MDL		LOQ / CI	RQL	Units
TARGETS								
58-89-9	gamma-BHC (Lindane)	0.037	U	0.037		(0.50	ug/L
76-44-8	Heptachlor	0.027	U	0.027		(0.50	ug/L
1024-57-3	Heptachlor epoxide	0.096	U	0.096		(0.50	ug/L
72-20-8	Endrin	0.032	U	0.032		(0.50	ug/L
72-43-5	Methoxychlor	0.11	U	0.11		(0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70			10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88		4	5.00	ug/L
SURROGATES								
SURROGATES 2051-24-3	Decachlorobiphenyl	19.6		30 (57)	- 150 (171)	9	98%	SPK: 2

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates > 25% difference for detected

concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

- J = Estimated Value
- B = Analyte Found in Associated Method Blank
- N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration

was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit