

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: Q2177

SDG NO.: Q2177

Lab File ID: BF142747.D

DFTPP Injection Date: 06/12/2025

Instrument ID: BNA_F

DFTPP Injection Time: 09:21

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	35.5
68	Less than 2.0% of mass 69	0.5 (1.8) 1
69	Mass 69 relative abundance	34.0
70	Less than 2.0% of mass 69	0.1 (0.2) 1
127	10.0 - 80.0% of mass 198	48.0
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	28.1
365	Greater than 1% of mass 198	4.0
441	Present, but less than mass 443	17.8
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.5 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142748.D	06/12/2025	09:50
PB168259BS	PB168259BS	BF142750.D	06/12/2025	11:25