

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

51

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH	Contract: PORTU6	PORTU6		
Lab Code: CHEM	SAS No.: Q2177	SDG NO.: Q2177		
Lab File ID: BP024904.D	DFTPP Injection Date:	06/11/2025		
Instrument ID: BNA_P	DFTPP Injection Time:	09:29		

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	29.9
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	34.6
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	45.8
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	30.8
365	Greater than 1% of mass 198	4.4
441	Present, but less than mass 443	13.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	16.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024905.D	06/11/2025	10:09
B-187-SB01DL	Q2177-02DL	BP024917.D	06/11/2025	18:27