

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG No.: Q2177

Instrument ID: BNA_M Calibration Date/Time: 06/06/2025 21:23

Lab File ID: BM050223.D Init. Calib. Date(s): 06/05/2025 06/05/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:20 13:56

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.361	1.385		1.8	
2-Fluorophenol	1.199	1.211		1.0	
Phenol-d6	1.578	1.509		-4.4	
1,4-Dichlorobenzene	1.534	1.456		-5.1	20.0
2-Methylphenol	1.102	1.030		-6.5	
3+4-Methylphenols	1.474	1.350		-8.4	
Nitrobenzene-d5	0.385	0.383		-0.5	
Hexachloroethane	0.555	0.553		-0.4	
Nitrobenzene	0.350	0.352		0.6	
Hexachlorobutadiene	0.190	0.178		-6.3	20.0
2,4,6-Trichlorophenol	0.380	0.368		-3.2	20.0
2-Fluorobiphenyl	1.477	1.470		-0.5	
2,4,5-Trichlorophenol	0.417	0.402		-3.6	
2,4-Dinitrotoluene	0.360	0.337		-6.4	
2,4,6-Tribromophenol	0.227	0.216		-4.8	
Hexachlorobenzene	0.249	0.244		-2.0	
Pentachlorophenol	0.162	0.221		36.4	20.0
Terphenyl-d14	1.055	1.019		-3.4	

All other compounds must meet a minimum RRF of 0.010.