

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

EPA Sample No.: SSTDICCC040 Date Analyzed: 06/05/2025

Lab File ID: BM050198.D Time Analyzed: 11:57

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	308806	7.793	1316350	10.58	819036	14.42
UPPER LIMIT	617612	8.293	2632700	11.075	1638070	14.916
LOWER LIMIT	154403	7.293	658175	10.075	409518	13.916
EPA SAMPLE NO.						
01 PB168224TB	259307	7.79	956559	10.58	522891	14.42

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.