

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177

EPA Sample No.: SSTDICCC040 Date Analyzed: 06/05/2025

Lab File ID: BM050198.D Time Analyzed: 11:57

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1590670	17.151	1484490	21.386	1410380	24.374
UPPER LIMIT	3181340	17.651	2968980	21.886	2820760	24.874
LOWER LIMIT	795335	16.651	742245	20.886	705190	23.874
EPA SAMPLE NO.						
01 PB168224TB	932766	17.15	869671	21.38	921150	24.37

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.