

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG No.: Q2177

Instrument ID: MSVOA_N Calibration Date/Time: 06/03/2025 11:30

Lab File ID: VN086835.D Init. Calib. Date(s): 05/16/2025 05/16/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 16:31 18:32

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.733	0.619		-15.55	20
1,1-Dichloroethene	0.558	0.513		-8.06	20
2-Butanone	0.399	0.374		-6.27	20
Carbon Tetrachloride	0.433	0.409		-5.54	20
Chloroform	1.119	1.032		-7.78	20
Benzene	1.433	1.362		-4.95	20
1,2-Dichloroethane	0.446	0.408		-8.52	20
Trichloroethene	0.346	0.320		-7.51	20
Tetrachloroethene	0.383	0.305		-20.37	20
Chlorobenzene	1.096	1.058	0.3	-3.47	20
1,2-Dichloroethane-d4	0.691	0.613		-11.29	20
Dibromofluoromethane	0.298	0.292		-2.01	20
Toluene-d8	1.234	1.163		-5.75	20
4-Bromofluorobenzene	0.436	0.409		-6.19	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.