

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2177

Client: Portal Partners Tri-Venture

Analytical Method: SW8260D

Datafile : VN086837.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0603WBS01	Vinyl chloride	20	16.7	ug/L	84			70 (65)	130 (117)	
	1,1-Dichloroethene	20	18.5	ug/L	93			70 (74)	130 (110)	
	2-Butanone	100	94.0	ug/L	94			40 (65)	160 (122)	
	Carbon Tetrachloride	20	18.8	ug/L	94			70 (77)	130 (113)	
	Chloroform	20	18.3	ug/L	92			70 (79)	130 (113)	
	Benzene	20	19.2	ug/L	96			70 (82)	130 (109)	
	1,2-Dichloroethane	20	18.3	ug/L	92			70 (80)	130 (115)	
	Trichloroethene	20	18.5	ug/L	93			70 (77)	130 (113)	
	Tetrachloroethene	20	15.8	ug/L	79			70 (67)	130 (123)	
	Chlorobenzene	20	19.5	ug/L	98			70 (82)	130 (109)	